

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100419\
 Data File : VW013440.D
 Acq On : 04 Oct 2019 15:45
 Operator : SY/VA
 Sample : K5213-04
 Misc : 5.08G/5ML/MSVOA W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S-2(5-10)

Quant Time: Oct 05 05:55:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100119S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 11:49:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	213801	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	307036	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	256265	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	120150	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	81352	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
35) Dibromofluoromethane	7.88	113	82094	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
50) Toluene-d8	10.32	98	351333	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
62) 4-Bromofluorobenzene	12.62	95	181012	81.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	162.16%	
Target Compounds						
4) Vinyl Chloride	2.35	62	11508	5.942	ug/l	97
16) Acetone	4.14	43	6905	14.111	ug/l	97
27) cis-1,2-Dichloroethene	7.17	96	88756	40.456	ug/l	89
64) Tetrachloroethene	10.86	164	11340	5.816	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW100419\
Data File : VW013440.D
Acq On : 04 Oct 2019 15:45
Operator : SY/VA
Sample : K5213-04
Misc : 5.08G/5ML/MSVOA W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S-2(5-10)

Quant Time: Oct 05 05:55:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100119S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 11:49:57 2019
Response via : Initial Calibration

