

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100618\
 Data File : VW005873.D
 Acq On : 05 Oct 2018 02:14
 Operator : SY/AP
 Sample : J5206-02
 Misc : 5.26G/5ML/MSVOA W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T-1A

Quant Time: Oct 05 15:17:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100418S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 04 07:08:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	94740	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	136073	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	36827	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	31530	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	68147	59.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.58%	
35) Dibromofluoromethane	7.88	113	56337	63.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	126.34%	
50) Toluene-d8	10.33	98	19136	5.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	10.80%	
62) 4-Bromofluorobenzene	12.62	95	37344	28.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	57.06%	
Target Compounds						
20) Methylene Chloride	4.91	84	16187	9.189	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW100618\
Data File : VW005873.D
Acq On : 05 Oct 2018 02:14
Operator : SY/AP
Sample : J5206-02
Misc : 5.26G/5ML/MSVOA W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T-1A

Quant Time: Oct 05 15:17:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100418S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 04 07:08:57 2018
Response via : Initial Calibration

