

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW100618\
 Data File : VW005867.D
 Acq On : 04 Oct 2018 23:13
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 05 01:17:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100418S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 04 07:08:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.286	0.315	-10.1	110	0.00
3 P	Chloromethane	0.344	0.343	0.3	106	0.00
4 C	Vinyl Chloride	0.504	0.499	1.0#	100	0.00
5 T	Bromomethane	0.351	0.330	6.0	96	0.00
6 T	Chloroethane	0.308	0.302	1.9	97	0.00
7 T	Trichlorofluoromethane	0.347	0.327	5.8	91	0.00
8 T	Diethyl Ether	0.263	0.259	1.5	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.479	0.492	-2.7	103	-0.01
10 T	Methyl Iodide	0.778	0.790	-1.5	97	0.00
11 T	Tert butyl alcohol	0.037	0.035	5.4	101	-0.01
12 CM	1,1-Dichloroethene	0.470	0.484	-3.0#	101	0.00
13 T	Acrolein	0.037	0.042	-13.5	122	0.00
14 T	Allyl chloride	0.762	0.797	-4.6	100	0.00
15 T	Acrylonitrile	0.111	0.120	-8.1	99	0.00
16 T	Acetone	0.101	0.119	-17.8	115	0.00
17 T	Carbon Disulfide	1.485	1.553	-4.6	99	0.00
18 T	Methyl Acetate	0.288	0.308	-6.9	101	0.00
19 T	Methyl tert-butyl Ether	0.737	0.765	-3.8	92	0.00
20 T	Methylene Chloride	0.721	0.549	23.9#	92	0.00
21 T	trans-1,2-Dichloroethene	0.513	0.542	-5.7	100	0.00
22 T	Diisopropyl ether	1.496	1.490	0.4	90	0.00
23 T	Vinyl Acetate	0.850	0.911	-7.2	94	0.00
24 P	1,1-Dichloroethane	0.970	0.945	2.6	94	0.00
25 T	2-Butanone	0.142	0.158	-11.3	105	0.00
26 T	2,2-Dichloropropane	0.606	0.589	2.8	99	0.00
27 T	cis-1,2-Dichloroethene	0.571	0.577	-1.1	97	0.00
28 T	Bromochloromethane	0.343	0.320	6.7	93	0.00
29 T	Tetrahydrofuran	0.088	0.100	-13.6	101	0.00
30 C	Chloroform	1.040	1.005	3.4#	93	0.00
31 T	Cyclohexane	0.830	0.834	-0.5	98	0.00
32 T	1,1,1-Trichloroethane	0.888	0.850	4.3	94	0.00
33 S	1,2-Dichloroethane-d4	0.607	0.579	4.6	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.328	0.343	-4.6	95	0.00
36 T	1,1-Dichloropropene	0.497	0.553	-11.3	98	0.00
37 T	Ethyl Acetate	0.207	0.228	-10.1	93	0.00
38 T	Carbon Tetrachloride	0.554	0.584	-5.4	97	0.00
39 T	Methylcyclohexane	0.560	0.628	-12.1	97	0.00
40 TM	Benzene	1.440	1.475	-2.4	91	0.00
41 T	Methacrylonitrile	0.124	0.141	-13.7	90	0.00
42 TM	1,2-Dichloroethane	0.484	0.497	-2.7	89	0.00
43 T	Isopropyl Acetate	0.409	0.448	-9.5	92	0.00
44 TM	Trichloroethene	0.381	0.382	-0.3	91	0.00
45 C	1,2-Dichloropropane	0.341	0.339	0.6#	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW100618\
 Data File : VW005867.D
 Acq On : 04 Oct 2018 23:13
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 05 01:17:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100418S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 04 07:08:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.187	0.189	-1.1	89	0.00
47 T	Bromodichloromethane	0.494	0.511	-3.4	93	0.00
48 T	Methyl methacrylate	0.198	0.215	-8.6	89	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	93	-0.01
50 S	Toluene-d8	1.301	1.417	-8.9	99	0.00
51 T	4-Methyl-2-Pentanone	0.213	0.250	-17.4	98	0.00
52 CM	Toluene	0.918	1.022	-11.3#	98	0.00
53 T	t-1,3-Dichloropropene	0.482	0.544	-12.9	98	0.00
54 T	cis-1,3-Dichloropropene	0.540	0.611	-13.1	98	0.00
55 T	1,1,2-Trichloroethane	0.269	0.294	-9.3	97	0.00
56 T	Ethyl methacrylate	0.323	0.388	-20.1#	99	0.00
57 T	1,3-Dichloropropane	0.459	0.510	-11.1	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.133	0.155	-16.5	100	0.00
59 T	2-Hexanone	0.142	0.178	-25.4#	102	0.00
60 T	Dibromochloromethane	0.336	0.365	-8.6	94	0.00
61 T	1,2-Dibromoethane	0.250	0.274	-9.6	95	0.00
62 S	4-Bromofluorobenzene	0.481	0.532	-10.6	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.390	0.394	-1.0	100	0.00
65 PM	Chlorobenzene	1.093	1.092	0.1	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.405	0.398	1.7	93	0.00
67 C	Ethyl Benzene	1.872	1.972	-5.3#	98	0.00
68 T	m/p-Xylenes	0.721	0.761	-5.5	97	0.00
69 T	o-Xylene	0.670	0.708	-5.7	97	0.00
70 T	Styrene	1.137	1.218	-7.1	98	0.00
71 P	Bromoform	0.223	0.231	-3.6	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.440	3.671	-6.7	99	0.00
74 T	N-amyl acetate	0.749	0.828	-10.5	100	0.00
75 P	1,1,2,2-Tetrachloroethane	0.590	0.607	-2.9	96	0.00
76 T	1,2,3-Trichloropropane	0.454	0.424	6.6	94	0.00
77 T	Bromobenzene	0.849	0.860	-1.3	96	0.00
78 T	n-propylbenzene	4.092	4.284	-4.7	97	0.00
79 T	2-Chlorotoluene	2.437	2.458	-0.9	95	0.00
80 T	1,3,5-Trimethylbenzene	2.988	3.115	-4.3	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.180	0.204	-13.3	103	0.00
82 T	4-Chlorotoluene	2.569	2.627	-2.3	97	0.00
83 T	tert-Butylbenzene	2.492	2.655	-6.5	98	0.00
84 T	1,2,4-Trimethylbenzene	3.068	3.146	-2.5	93	0.00
85 T	sec-Butylbenzene	3.597	3.859	-7.3	101	0.00
86 T	p-Isopropyltoluene	3.228	3.521	-9.1	100	0.00
87 T	1,3-Dichlorobenzene	1.713	1.735	-1.3	97	0.00
88 T	1,4-Dichlorobenzene	1.710	1.736	-1.5	99	0.00
89 T	n-Butylbenzene	2.973	3.250	-9.3	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW100618\
Data File : VW005867.D
Acq On : 04 Oct 2018 23:13
Operator : SY/AP
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Oct 05 01:17:32 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100418S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 04 07:08:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.607	0.612	-0.8	99	0.00
91 T	1,2-Dichlorobenzene	1.541	1.560	-1.2	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.114	0.121	-6.1	101	0.00
93 T	1,2,4-Trichlorobenzene	1.077	1.150	-6.8	101	0.00
94 T	Hexachlorobutadiene	0.681	0.718	-5.4	104	0.00
95 T	Naphthalene	1.855	2.122	-14.4	101	0.00
96 T	1,2,3-Trichlorobenzene	0.968	1.038	-7.2	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6