

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100621\
 Data File : VW020312.D
 Acq On : 06 Oct 2021 13:11
 Operator : SY/VA
 Sample : M3973-22
 Misc : 8.16g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Oct 07 01:51:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 07 01:49:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	225856	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	151520	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.438	152	40578	25.000	ug/L	-0.12
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	78440	25.199	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.800%
7) Chloroethane-d5	2.885	69	30539	22.642	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.560%
11) 1,1-Dichloroethene-d2	4.019	63	97739	15.247	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	61.000%
21) 2-Butanone-d5	7.086	46	42065	53.197	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.400%
24) Chloroform-d	7.653	84	140860	21.720	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.880%
26) 1,2-Dichloroethane-d4	8.311	65	80935	22.804	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.200%
32) Benzene-d6	8.281	84	279790	30.498	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	122.000%
36) 1,2-Dichloropropane-d6	9.280	67	120567	41.434	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	165.720%#
41) Toluene-d8	10.323	98	272006	32.736	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	130.960%#
43) trans-1,3-Dichloroprop...	10.579	79	33567	28.351	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	113.400%
47) 2-Hexanone-d5	10.926	63	31588	78.541	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	157.080%#
56) 1,1,2,2-Tetrachloroeth...	12.694	84	51767	25.489	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	101.960%
66) 1,2-Dichlorobenzene-d4	13.859	152	54541	38.384	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	153.520%#
Target Compounds						Qvalue
5) Vinyl chloride	2.361	62	18882	4.588	ug/L #	76
12) 1,1-Dichloroethene	4.025	96	1002	0.319	ug/L #	1
13) Acetone	4.135	43	10378	12.766	ug/L	94
14) Carbon disulfide	4.385	76	3604	0.349	ug/L	97
16) Methylene chloride	4.921	84	6220	1.542	ug/L	91
19) 1,1-Dichloroethane	6.226	63	5122	0.783	ug/L	94
20) cis-1,2-Dichloroethene	7.171	96	5324	1.582	ug/L	100
22) 2-Butanone	7.177	43	3269	3.271	ug/L	80
29) Cyclohexane	7.957	56	271088	60.178	ug/L	91
33) Benzene	8.329	78	72771	6.906	ug/L	100
34) Trichloroethene	9.122	95	4736	1.827	ug/L #	20
35) Methylcyclohexane	9.335	83	1072869	237.271	ug/L	92
37) 1,2-Dichloropropane	9.122	63	1571	0.598	ug/L	98
38) Bromodichloromethane	9.616	83	17175	5.370	ug/L #	35
40) 4-Methyl-2-pentanone	10.213	43	4888	3.409	ug/L #	51
42) Toluene	10.390	91	28315	2.672	ug/L	95
45) 1,1,2-Trichloroethane	10.762	97	232941	140.266	ug/L #	21
48) 2-Hexanone	10.939	43	514881	492.180	ug/L #	29
50) 1,2-Dibromoethane	11.225	107	774	0.507	ug/L #	1

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100621\
 Data File : VW020312.D
 Acq On : 06 Oct 2021 13:11
 Operator : SY/VA
 Sample : M3973-22
 Misc : 8.16g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Oct 07 01:51:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 07 01:49:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


52) Ethylbenzene	11.731	91	37903	3.276	ug/L	91
53) m,p-Xylene	11.847	106	30607	7.161	ug/L	90
54) o-Xylene	12.170	106	134254	33.543	ug/L	99
55) Styrene	12.170	104	7382	1.072	ug/L #	1
57) 1,1,2,2-Tetrachloroethane	12.694	83	38135	19.314	ug/L #	72
60) Isopropylbenzene	12.463	105	321631	53.392	ug/L	98
61) 1,2,3-Trichloropropane	12.804	75	3132	3.894	ug/L #	1
62) 1,3,5-Trimethylbenzene	12.944	105	229697	47.127	ug/L	99
63) 1,2,4-Trimethylbenzene	13.249	105	696054	142.898	ug/L	100
65) 1,4-Dichlorobenzene	13.572	146	1840	0.719	ug/L #	1
68) 1,2-Dibromo-3-chloropr...	14.499	75	10343	60.589	ug/L #	37
69) 1,3,5-Trichlorobenzene	14.633	180	858	0.477	ug/L #	73
70) 1,2,4-trichlorobenzene	15.127	180	6111	4.353	ug/L #	1
71) Naphthalene	15.365	128	818139	334.542	ug/L	96
72) 1,2,3-Trichlorobenzene	15.578	180	18906	15.417	ug/L #	67


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100621\
 Data File : VW020312.D
 Acq On : 06 Oct 2021 13:11
 Operator : SY/VA
 Sample : M3973-22
 Misc : 8.16g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Oct 07 01:51:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 07 01:49:38 2021
 Response via : Initial Calibration









#5
 Vinyl chloride
 Concen: 4.588 ug/L
 RT: 2.361 min Scan# 75
 Delta R.T. -0.000 min
 Lab File: VW020312.D
 Acq: 06 Oct 2021 13:11




Instrument :
 MSVOA_W
 ClientSampled :


Tgt Ion: 62 Resp: 18882
 Ion Ratio Lower Upper
 62 100
 64 49.1 24.6 45.6#



#12
 1,1-Dichloroethene
 Concen: 0.319 ug/L
 RT: 4.025 min Scan# 348
 Delta R.T. -0.018 min
 Lab File: VW020312.D
 Acq: 06 Oct 2021 13:11

Tgt Ion: 96 Resp: 1002
 Ion Ratio Lower Upper
 96 100
 61 1278.9 133.9 248.7#
 63 9429.1 77.3 143.7#









#13
Acetone
Concen: 12.766 ug/L
RT: 4.135 min Scan# 366
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 43 Resp: 10378
Ion Ratio Lower Upper
43 100
58 25.8 0.0 58.4



#14
Carbon disulfide
Concen: 0.349 ug/L
RT: 4.385 min Scan# 407
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Tgt Ion: 76 Resp: 3604
Ion Ratio Lower Upper
76 100
78 8.3 7.4 11.2









#16
Methylene chloride
Concen: 1.542 ug/L
RT: 4.921 min Scan# 495
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 84 Resp: 6220
Ion Ratio Lower Upper
84 100
86 74.3 44.7 83.1
49 136.8 101.7 188.9



#19
1,1-Dichloroethane
Concen: 0.783 ug/L
RT: 6.226 min Scan# 709
Delta R.T. 0.006 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Tgt Ion: 63 Resp: 5122
Ion Ratio Lower Upper
63 100
65 35.4 21.9 40.7
83 14.4 9.0 16.6









#20
 cis-1,2-Dichloroethene
 Concen: 1.582 ug/L
 RT: 7.171 min Scan# 864
 Delta R.T. -0.000 min
 Lab File: VW020312.D
 Acq: 06 Oct 2021 13:11




Instrument :
 MSVOA_W
 ClientSampled :


Tgt Ion: 96 Resp: 5324
 Ion Ratio Lower Upper
 96 100
 61 143.3 100.3 186.3
 68 0.0 0.0 0.0



#22
 2-Butanone
 Concen: 3.271 ug/L
 RT: 7.177 min Scan# 865
 Delta R.T. -0.000 min
 Lab File: VW020312.D
 Acq: 06 Oct 2021 13:11

Tgt Ion: 43 Resp: 3269
 Ion Ratio Lower Upper
 43 100
 72 14.4 12.2 36.6









#29
Cyclohexane
Concen: 60.178 ug/L
RT: 7.957 min Scan# 993
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 56 Resp: 271088
Ion Ratio Lower Upper
56 100
69 28.9 23.2 34.8
84 68.3 63.7 95.5



#33
Benzene
Concen: 6.906 ug/L
RT: 8.329 min Scan# 1054
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Tgt Ion: 78 Resp: 72771









#34
Trichloroethene
Concen: 1.827 ug/L
RT: 9.122 min Scan# 1184
Delta R.T. 0.030 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Instrument :
MSVOA_W
ClientSampleId :

Tgt Ion: 95 Resp: 4736




Ion	Ratio	Lower	Upper
95	100		
97	36.5	43.2	80.2#
132	0.0	62.3	115.7#
130	0.0	66.1	122.7#




#35
Methylcyclohexane
Concen: 237.271 ug/L
RT: 9.335 min Scan# 1219
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Tgt Ion: 83 Resp: 1072869

Ion	Ratio	Lower	Upper
83	100		
55	95.1	69.7	104.5
98	43.8	38.2	57.2









#37
1,2-Dichloropropane
Concen: 0.598 ug/L
RT: 9.122 min Scan# 1184
Delta R.T. -0.250 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 63 Resp: 1571
Ion Ratio Lower Upper
63 100
112 3.8 3.5 5.3



#38
Bromodichloromethane
Concen: 5.370 ug/L
RT: 9.616 min Scan# 1265
Delta R.T. -0.031 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Tgt Ion: 83 Resp: 17175
Ion Ratio Lower Upper
83 100
85 7.7 43.4 80.6#
127 0.0 6.2 9.2#









#40
4-Methyl-2-pentanone
Concen: 3.409 ug/L
RT: 10.213 min Scan# 1363
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 43 Resp: 4888
Ion Ratio Lower Upper
43 100
58 4.0 29.0 43.4#
100 0.0 10.2 15.4#



#42
Toluene
Concen: 2.672 ug/L
RT: 10.390 min Scan# 1392
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11


Tgt Ion: 91 Resp: 28315
Ion Ratio Lower Upper
91 100
92 55.3 41.5 77.1








#45
1,1,2-Trichloroethane
Concen: 140.266 ug/L
RT: 10.762 min Scan# 1453
Delta R.T. -0.025 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11


Instrument :
MSVOA_W
ClientSampled :





Tgt Ion: 97 Resp: 232941
Ion Ratio Lower Upper
97 100
99 0.2 42.8 79.4#
83 9.3 60.9 113.1#
85 3.9 39.6 73.6#




#48
2-Hexanone
Concen: 492.180 ug/L
RT: 10.939 min Scan# 1482
Delta R.T. -0.031 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11




Tgt Ion: 43 Resp: 514881
Ion Ratio Lower Upper
43 100
58 3.1 39.5 59.3#
57 64.9 14.2 21.4#
100 0.1 8.2 12.4#








#50
1,2-Dibromoethane
Concen: 0.507 ug/L
RT: 11.225 min Scan# 1529
Delta R.T. -0.012 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Instrument :
MSVOA_W
ClientSampled :







Tgt Ion:107 Resp: 774
Ion Ratio Lower Upper
107 100
109 768.1 64.5 119.9#
188 0.0 3.2 4.8#



#52
Ethylbenzene
Concen: 3.276 ug/L
RT: 11.731 min Scan# 1612
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Tgt Ion: 91 Resp: 37903
Ion Ratio Lower Upper
91 100
106 25.6 21.3 39.5









#53
m,p-Xylene
Concen: 7.161 ug/L
RT: 11.847 min Scan# 1631
Delta R.T. 0.006 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:106 Resp: 30607
Ion Ratio Lower Upper
106 100
91 227.5 148.2 275.2



#54
o-Xylene
Concen: 33.543 ug/L
RT: 12.170 min Scan# 1684
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Tgt Ion:106 Resp: 134254
Ion Ratio Lower Upper
106 100
91 225.1 156.1 289.9





#55

Styrene

Concen: 1.072 ug/L

RT: 12.170 min Scan# 1684

Delta R.T. -0.012 min

Lab File: VW020312.D

Acq: 06 Oct 2021 13:11

Tgt Ion:104 Resp: 7382

Ion Ratio Lower Upper





104 100

78 345.5 33.2 61.7#

Instrument :

MSVOA_W

ClientSampled :



#57

1,1,2,2-Tetrachloroethane

Concen: 19.314 ug/L

RT: 12.694 min Scan# 1770

Delta R.T. -0.018 min

Lab File: VW020312.D

Acq: 06 Oct 2021 13:11




Tgt Ion: 83 Resp: 38135


Ion Ratio Lower Upper

83 100

85 42.9 45.9 85.3#

131 1.0 7.4 11.2#









#60
Isopropylbenzene
Concen: 53.392 ug/L
RT: 12.463 min Scan# 1732
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:105 Resp: 321631
Ion Ratio Lower Upper
105 100
120 25.5 20.3 30.5
77 19.1 13.8 20.6



#61
1,2,3-Trichloropropane
Concen: 3.894 ug/L
RT: 12.804 min Scan# 1788
Delta R.T. 0.036 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Tgt Ion: 75 Resp: 3132
Ion Ratio Lower Upper
75 100
77 1400.7 25.1 37.7#
110 942.9 29.4 44.0#









#62
1,3,5-Trimethylbenzene
Concen: 47.127 ug/L
RT: 12.944 min Scan# 1811
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:105 Resp: 229697
Ion Ratio Lower Upper
105 100
120 47.8 37.9 56.9



#63
1,2,4-Trimethylbenzene
Concen: 142.898 ug/L
RT: 13.249 min Scan# 1861
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Tgt Ion:105 Resp: 696054
Ion Ratio Lower Upper
105 100
120 44.2 35.4 53.2









#65
1,4-Dichlorobenzene
Concen: 0.719 ug/L
RT: 13.572 min Scan# 1914
Delta R.T. -0.006 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:146 Resp: 1840
Ion Ratio Lower Upper
146 100
111 9137.4 30.9 57.5#
148 28.1 46.1 85.7#



#68
1,2-Dibromo-3-chloropropane
Concen: 60.589 ug/L
RT: 14.499 min Scan# 2066
Delta R.T. 0.018 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11


Tgt Ion: 75 Resp: 10343
Ion Ratio Lower Upper
75 100
155 47.1 51.4 77.0#
157 1.9 70.8 106.2#








#69
1,3,5-Trichlorobenzene
Concen: 0.477 ug/L
RT: 14.633 min Scan# 2088
Delta R.T. 0.006 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Instrument :
MSVOA_W
ClientSampleId :







Tgt Ion:180 Resp: 858
Ion Ratio Lower Upper
180 100
182 99.3 73.4 110.0
184 0.0 24.2 36.4#
145 0.0 23.7 35.5#



#70
1,2,4-trichlorobenzene
Concen: 4.353 ug/L
RT: 15.127 min Scan# 2169
Delta R.T. -0.006 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Tgt Ion:180 Resp: 6111
Ion Ratio Lower Upper
180 100
182 572.1 75.8 113.8#
145 6885.8 24.8 37.2#









#71
Naphthalene
Concen: 334.542 ug/L
RT: 15.365 min Scan# 2208
Delta R.T. -0.000 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Instrument :
MSVOA_W
ClientSampled :

Tgt Ion:128 Resp: 818139
Ion Ratio Lower Upper
128 100
127 11.9 10.5 15.7
129 9.4 8.7 13.1



#72
1,2,3-Trichlorobenzene
Concen: 15.417 ug/L
RT: 15.578 min Scan# 2243
Delta R.T. 0.024 min
Lab File: VW020312.D
Acq: 06 Oct 2021 13:11

Tgt Ion:180 Resp: 18906
Ion Ratio Lower Upper
180 100
182 70.0 74.8 112.2#
145 0.0 26.2 39.4#

