

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100621\
 Data File : VW020311.D
 Acq On : 06 Oct 2021 12:47
 Operator : SY/VA
 Sample : VW1006SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK452

Quant Time: Oct 07 03:11:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 07 01:49:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	211519	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	198698	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	94294	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	74076	25.410	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 101.640%			
7) Chloroethane-d5	2.885	69	31851	25.215	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 100.880%			
11) 1,1-Dichloroethene-d2	4.019	63	103000	17.157	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery = 68.640%			
21) 2-Butanone-d5	7.086	46	40861	55.177	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 110.360%			
24) Chloroform-d	7.653	84	144970	23.869	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery = 95.480%			
26) 1,2-Dichloroethane-d4	8.311	65	84582	25.447	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery = 101.800%			
32) Benzene-d6	8.281	84	285476	23.729	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery = 94.920%			
36) 1,2-Dichloropropane-d6	9.274	67	92344	24.200	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery = 96.800%			
41) Toluene-d8	10.323	98	256125	23.506	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery = 94.040%			
43) trans-1,3-Dichloroprop...	10.579	79	37402	24.089	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery = 96.360%			
47) 2-Hexanone-d5	10.927	63	28131	53.338	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 106.680%			
56) 1,1,2,2-Tetrachloroeth...	12.688	84	67312	25.274	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery = 101.080%			
66) 1,2-Dichlorobenzene-d4	13.853	152	82158	24.882	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery = 99.520%			
Target Compounds						Qvalue
16) Methylene chloride	4.922	84	8309	2.200	ug/L	99
42) Toluene	10.390	91	4587	0.330	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100621\
Data File : VW020311.D
Acq On : 06 Oct 2021 12:47
Operator : SY/VA
Sample : VW1006SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK452

Quant Time: Oct 07 03:11:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 07 01:49:38 2021
Response via : Initial Calibration

