

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100621\
 Data File : VW020328.D
 Acq On : 06 Oct 2021 21:23
 Operator : SY/VA
 Sample : M4000-01
 Misc : 7.02g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW8Z6

Quant Time: Oct 07 01:55:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 07 01:49:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	313584	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	240216	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	81612	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	88870	20.563	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.240%
7) Chloroethane-d5	2.879	69	43439	23.196	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.800%
11) 1,1-Dichloroethene-d2	4.019	63	121207	13.619	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	54.480%
21) 2-Butanone-d5	7.086	46	12900	11.750	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	23.500%
24) Chloroform-d	7.653	84	225163	25.007	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.040%
26) 1,2-Dichloroethane-d4	8.311	65	132837	26.957	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	107.840%
32) Benzene-d6	8.274	84	392017	26.953	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	107.800%
36) 1,2-Dichloropropane-d6	9.274	67	141765	30.730	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	122.920%#
41) Toluene-d8	10.323	98	292512	22.205	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.840%
43) trans-1,3-Dichloroprop...	10.579	79	23862	12.712	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	50.840%
47) 2-Hexanone-d5	10.926	63	7810	12.249	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	24.500%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	90028	27.961	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	111.840%
66) 1,2-Dichlorobenzene-d4	13.853	152	69912	24.464	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.840%
Target Compounds						Qvalue
15) Methyl Acetate	4.677	43	1639	0.795	ug/L	# 89
16) Methylene chloride	4.915	84	8615	1.538	ug/L	96
34) Trichloroethene	9.091	95	3524	0.858	ug/L	99
42) Toluene	10.390	91	5357	0.319	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100621\
Data File : VW020328.D
Acq On : 06 Oct 2021 21:23
Operator : SY/VA
Sample : M4000-01
Misc : 7.02g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW8Z6

Quant Time: Oct 07 01:55:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 07 01:49:38 2021
Response via : Initial Calibration

