

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100621\
 Data File : VW020362.D
 Acq On : 07 Oct 2021 13:16
 Operator : SY/VA
 Sample : VIBLK461
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK461

Quant Time: Oct 08 05:58:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 08 05:56:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	257457	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	242722	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	121008	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	90604	25.534	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 102.120%	
7) Chloroethane-d5	2.879	69	43590	28.351	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 113.400%	
11) 1,1-Dichloroethene-d2	4.013	63	132638	18.152	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 72.600%	
21) 2-Butanone-d5	7.086	46	52992	58.790	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 117.580%	
24) Chloroform-d	7.653	84	180041	24.354	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 97.400%	
26) 1,2-Dichloroethane-d4	8.305	65	106138	26.235	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 104.920%	
32) Benzene-d6	8.275	84	367848	25.030	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 100.120%	
36) 1,2-Dichloropropane-d6	9.274	67	116215	24.931	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 99.720%	
41) Toluene-d8	10.323	98	328631	24.690	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 98.760%	
43) trans-1,3-Dichloroprop...	10.579	79	46954	24.756	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 99.040%	
47) 2-Hexanone-d5	10.927	63	37050	57.507	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 115.020%	
56) 1,1,2,2-Tetrachloroeth...	12.695	84	85920	26.410	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 105.640%	
66) 1,2-Dichlorobenzene-d4	13.853	152	109044	25.734	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 102.920%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100621\
Data File : VW020362.D
Acq On : 07 Oct 2021 13:16
Operator : SY/VA
Sample : VIBLK461
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK461

Quant Time: Oct 08 05:58:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 08 05:56:34 2021
Response via : Initial Calibration

