

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100621\
 Data File : VW020368.D
 Acq On : 07 Oct 2021 15:40
 Operator : SY/VA
 Sample : M4000-05
 Misc : 7.26g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW900

Quant Time: Oct 08 05:58:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 08 05:56:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	220578	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	137706	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	35102	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	100074	32.919	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	131.680%
7) Chloroethane-d5	2.879	69	59792	45.391	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	181.560%#
11) 1,1-Dichloroethene-d2	4.013	63	108538	17.337	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.360%
21) 2-Butanone-d5	7.086	46	69623	90.154	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	180.300%#
24) Chloroform-d	7.653	84	180589	28.513	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	114.040%
26) 1,2-Dichloroethane-d4	8.311	65	114623	33.069	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	132.280%#
32) Benzene-d6	8.275	84	301428	36.153	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	144.600%#
36) 1,2-Dichloropropane-d6	9.274	67	102391	38.717	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	154.880%#
41) Toluene-d8	10.323	98	172664	22.865	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.440%
43) trans-1,3-Dichloroprop...	10.579	79	26098	24.254	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	97.000%
47) 2-Hexanone-d5	10.927	63	32051	87.686	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	175.380%#
56) 1,1,2,2-Tetrachloroeth...	12.695	84	53997	29.255	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	117.000%
66) 1,2-Dichlorobenzene-d4	13.853	152	27359	22.258	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.040%
Target Compounds					Qvalue	
13) Acetone	4.141	43	8827	11.118	ug/L	75
14) Carbon disulfide	4.379	76	10768	1.067	ug/L #	94
15) Methyl Acetate	4.690	43	2177	1.500	ug/L #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100621\
Data File : VW020368.D
Acq On : 07 Oct 2021 15:40
Operator : SY/VA
Sample : M4000-05
Misc : 7.26g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW900

Quant Time: Oct 08 05:58:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 08 05:56:34 2021
Response via : Initial Calibration

