

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100621\
 Data File : VW020311.D
 Acq On : 06 Oct 2021 12:47
 Operator : SY/VA
 Sample : VW1006SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK452

Quant Time: Oct 07 03:11:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 07 01:49:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	211519	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	198698	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	94294	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	74076	25.410	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.640%
7) Chloroethane-d5	2.885	69	31851	25.215	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.880%
11) 1,1-Dichloroethene-d2	4.019	63	103000	17.157	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.640%
21) 2-Butanone-d5	7.086	46	40861	55.177	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.360%
24) Chloroform-d	7.653	84	144970	23.869	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.480%
26) 1,2-Dichloroethane-d4	8.311	65	84582	25.447	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	101.800%
32) Benzene-d6	8.281	84	285476	23.729	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.920%
36) 1,2-Dichloropropane-d6	9.274	67	92344	24.200	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.800%
41) Toluene-d8	10.323	98	256125	23.506	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.040%
43) trans-1,3-Dichloroprop...	10.579	79	37402	24.089	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	96.360%
47) 2-Hexanone-d5	10.927	63	28131	53.338	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.680%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	67312	25.274	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	101.080%
66) 1,2-Dichlorobenzene-d4	13.853	152	82158	24.882	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.520%
Target Compounds						Qvalue
16) Methylene chloride	4.922	84	8309	2.200	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100621\
Data File : VW020311.D
Acq On : 06 Oct 2021 12:47
Operator : SY/VA
Sample : VW1006SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK452

Quant Time: Oct 07 03:11:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 07 01:49:38 2021
Response via : Initial Calibration





#16
Methylene chloride
Concen: 2.200 ug/L
RT: 4.922 min Scan# 495
Delta R.T. -0.000 min
Lab File: VW020311.D
Acq: 06 Oct 2021 12:47

Instrument :
MSVOA_W
ClientSampled :
VBLK452

Tgt Ion: 84 Resp: 8309
Ion Ratio Lower Upper
84 100
86 63.1 44.7 83.1
49 143.4 101.7 188.9

