

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100720\
 Data File : VW016806.D
 Acq On : 06 Oct 2020 20:45
 Operator : SY/VA
 Sample : L4235-08RE
 Misc : 4.72G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AK4RE

Quant Time: Oct 07 04:29:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM100720S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 07 03:05:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	566973	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	380900	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	87323	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	157474	28.15	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	112.60%
7) Chloroethane-d5	2.90	69	130271	27.76	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	111.04%
10) 1,1-Dichloroethene-d2	4.03	63	231021	19.45	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	77.80%
20) 2-Butanone-d5	7.09	46	35604	28.10	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	56.20%
24) Chloroform-d	7.65	84	294936	24.65	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.60%
26) 1,2-Dichloroethane-d4	8.31	65	139537	24.15	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.60%
29) Benzene-d6	8.27	84	603643	34.89	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	139.56%#
33) 1,2-Dichloropropane-d6	9.28	67	158121	34.01	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	136.04%#
37) Toluene-d8	10.33	98	504258	30.89	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	123.56%
38) trans-1,3-Dichloropropene-	10.59	79	50767	24.76	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	99.04%
39) 2-Hexanone-d5	10.93	63	27020	37.58	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.16%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	76998	20.50	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.00%
61) 1,2-Dichlorobenzene-d4	13.86	152	66727	24.09	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.36%

Target Compounds					Ovalue
13) Acetone	4.13	43	8265	8.621 ug/L	96
16) Methylene chloride	4.93	84	14936	1.909 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW100720\
Data File : VW016806.D
Acq On : 06 Oct 2020 20:45
Operator : SY/VA
Sample : L4235-08RE
Misc : 4.72G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AK4RE

Quant Time: Oct 07 04:29:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM100720S.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 07 03:05:19 2020
Response via : Initial Calibration

