

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100720\
 Data File : VW016812.D
 Acq On : 06 Oct 2020 23:05
 Operator : SY/VA
 Sample : L4209-08
 Misc : 5.85G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JLWT1

Quant Time: Oct 07 04:52:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM100720S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 07 03:05:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	560677	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	465845	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	193408	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	164805	29.80	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	119.20%
7) Chloroethane-d5	2.90	69	135405	29.18	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	116.72%
10) 1,1-Dichloroethene-d2	4.03	63	232153	19.77	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.08%
20) 2-Butanone-d5	7.08	46	61291	48.92	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.84%
24) Chloroform-d	7.65	84	329436	27.85	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	111.40%
26) 1,2-Dichloroethane-d4	8.31	65	165131	28.90	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	115.60%
29) Benzene-d6	8.28	84	673562	31.84	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	127.36%
33) 1,2-Dichloropropane-d6	9.28	67	181306	31.89	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	127.56%#
37) Toluene-d8	10.32	98	619376	31.02	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	124.08%
38) trans-1,3-Dichloropropene-	10.58	79	70884	28.27	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	113.08%
39) 2-Hexanone-d5	10.93	63	49149	55.89	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	111.78%
48) 1,1,1,2,2-Tetrachloroethane-	12.69	84	123965	26.99	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	107.96%
61) 1,2-Dichlorobenzene-d4	13.85	152	180812	29.47	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	117.88%

Target Compounds

					Ovalue
13) Acetone	4.13	43	10315	10.880	ug/L 98
16) Methylene chloride	4.93	84	12424	1.606	ug/L 95
35) Trichloroethene	9.10	95	1517	0.214	ug/L 94
43) 4-Methyl-2-pentanone	10.21	43	2413	0.847	ug/L 96
47) Tetrachloroethene	10.87	164	1182959	180.318	ug/L 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW100720\
Data File : VW016812.D
Acq On : 06 Oct 2020 23:05
Operator : SY/VA
Sample : L4209-08
Misc : 5.85G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JLWT1

Quant Time: Oct 07 04:52:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM100720S.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 07 03:05:19 2020
Response via : Initial Calibration

