

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100720\
 Data File : VW016815.D
 Acq On : 07 Oct 2020 00:16
 Operator : SY/VA
 Sample : L4254-06
 Misc : 5.68G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG3T3

Quant Time: Oct 07 05:14:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM100720S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 07 03:05:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	567053	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	480595	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	184216	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	158843	28.40	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	113.60%
7) Chloroethane-d5	2.89	69	131152	27.95	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	111.80%
10) 1,1-Dichloroethene-d2	4.03	63	218851	18.42	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.68%
20) 2-Butanone-d5	7.08	46	57012	45.00	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.00%
24) Chloroform-d	7.65	84	315437	26.36	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	105.44%
26) 1,2-Dichloroethane-d4	8.31	65	147675	25.55	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	102.20%
29) Benzene-d6	8.27	84	647788	29.68	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	118.72%
33) 1,2-Dichloropropane-d6	9.27	67	173879	29.64	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	118.56%
37) Toluene-d8	10.32	98	607355	29.49	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	117.96%
38) trans-1,3-Dichloropropene-	10.58	79	56347	21.78	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.12%
39) 2-Hexanone-d5	10.93	63	45389	50.03	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	100.06%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	114018	24.06	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.24%
61) 1,2-Dichlorobenzene-d4	13.85	152	170541	29.19	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	116.76%

Target Compounds

				Ovalue	
5) Vinyl chloride	2.37	62	7190	0.927 ug/L	# 19
13) Acetone	4.12	43	9748	10.166 ug/L	96
14) Carbon disulfide	4.39	76	19693	0.937 ug/L	98
16) Methylene chloride	4.92	84	10459	1.336 ug/L	94
47) Tetrachloroethene	10.87	164	36798	5.437 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW100720\
Data File : VW016815.D
Acq On : 07 Oct 2020 00:16
Operator : SY/VA
Sample : L4254-06
Misc : 5.68G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG3T3

Quant Time: Oct 07 05:14:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM100720S.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 07 03:05:19 2020
Response via : Initial Calibration

