

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100721\
 Data File : VW020377.D
 Acq On : 07 Oct 2021 22:36
 Operator : SY/VA
 Sample : M4070-04
 Misc : 6.09g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW7D5

Quant Time: Oct 08 07:39:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 08 07:38:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	312042	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	283636	25.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	122697	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	96369	22.41	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.64%
7) Chloroethane-d5	2.885	69	46643	25.03	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.12%
11) 1,1-Dichloroethene-d2	4.013	63	139338	15.73	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.92%
21) 2-Butanone-d5	7.080	46	48735	44.61	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.22%
24) Chloroform-d	7.653	84	195417	21.81	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.24%
26) 1,2-Dichloroethane-d4	8.311	65	108713	22.17	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.68%
32) Benzene-d6	8.274	84	389017	22.65	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.60%
36) 1,2-Dichloropropane-d6	9.274	67	125212	22.99	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	91.96%
41) Toluene-d8	10.323	98	349719	22.48	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.92%
43) trans-1,3-Dichloroprop...	10.579	79	46217	20.85	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.40%
47) 2-Hexanone-d5	10.926	63	32694	43.43	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.86%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	81421	21.42	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	85.68%
66) 1,2-Dichlorobenzene-d4	13.853	152	105304	24.51	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.04%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100721\
Data File : VW020377.D
Acq On : 07 Oct 2021 22:36
Operator : SY/VA
Sample : M4070-04
Misc : 6.09g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW7D5

Quant Time: Oct 08 07:39:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100621SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 08 07:38:51 2021
Response via : Initial Calibration

