

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW100818\
 Data File : VW005902.D
 Acq On : 08 Oct 2018 18:09
 Operator : SY/AP
 Sample : J5294-01
 Misc : 5.02G/5ML/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 HP-SUB

Quant Time: Oct 09 07:26:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 02:14:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	256736	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	388947	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	303910	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	102591	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	85660	34.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.74%	
35) Dibromofluoromethane	7.88	113	86339	34.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.54%	
50) Toluene-d8	10.33	98	266997	27.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	55.90%	
62) 4-Bromofluorobenzene	12.62	95	79629	21.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	43.82%	
Target Compounds						
20) Methylene Chloride	4.91	84	6758	2.71	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW100818\
Data File : VW005902.D
Acq On : 08 Oct 2018 18:09
Operator : SY/AP
Sample : J5294-01
Misc : 5.02G/5ML/MSVOA_W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
HP-SUB

Quant Time: Oct 09 07:26:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 02:14:04 2018
Response via : Initial Calibration

