

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW100818\
 Data File : VW005921.D
 Acq On : 09 Oct 2018 04:29
 Operator : SY/AP
 Sample : J5306-05
 Misc : 6.81G/5ML/MSVOA_W/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S-26-V

Quant Time: Oct 09 08:11:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 02:14:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	243188	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	376589	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	328376	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	154294	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	78801	33.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.76%	
35) Dibromofluoromethane	7.88	113	80857	33.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.26%	
50) Toluene-d8	10.33	98	261961	28.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	56.64%	
62) 4-Bromofluorobenzene	12.62	95	98994	28.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	56.26%	
Target Compounds						
16) Acetone	4.15	43	7210	15.29	ug/l	Qvalue 89
20) Methylene Chloride	4.92	84	19021	8.05	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW100818\
Data File : VW005921.D
Acq On : 09 Oct 2018 04:29
Operator : SY/AP
Sample : J5306-05
Misc : 6.81G/5ML/MSVOA_W/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S-26-V

Quant Time: Oct 09 08:11:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 02:14:04 2018
Response via : Initial Calibration

