

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW100818\
 Data File : VW005928.D
 Acq On : 09 Oct 2018 07:29
 Operator : SY/AP
 Sample : J5287-04
 Misc : 5.11G/5ML/MSVOA_W/SOIL
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TRACK-D-4

Quant Time: Oct 09 08:27:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 02:14:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	28158	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	45873	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	41640	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	17374	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	10840	39.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.32%	
35) Dibromofluoromethane	7.89	113	10534	35.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.94%	
50) Toluene-d8	10.33	98	34576	30.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	61.36%	
62) 4-Bromofluorobenzene	12.62	95	12720	29.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	59.34%	
Target Compounds						
16) Acetone	4.12	43	2607	47.75	ug/l	Qvalue # 88
20) Methylene Chloride	4.92	84	2362	8.63	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW100818\
Data File : VW005928.D
Acq On : 09 Oct 2018 07:29
Operator : SY/AP
Sample : J5287-04
Misc : 5.11G/5ML/MSVOA_W/SOIL
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TRACK-D-4

Quant Time: Oct 09 08:27:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 02:14:04 2018
Response via : Initial Calibration

