

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW100818\
 Data File : VW005908.D
 Acq On : 08 Oct 2018 21:10
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 6.68G/5ML/MSVOA_W/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 09 07:33:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 02:14:04 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	0.245	0.220	10.2	68	0.00
3 P	Chloromethane	0.288	0.268	6.9	76	0.00
4 C	Vinyl Chloride	0.458	0.469	-2.4#	77	0.00
5 T	Bromomethane	0.361	0.395	-9.4	82	0.00
6 T	Chloroethane	0.292	0.310	-6.2	80	0.00
7 T	Trichlorofluoromethane	0.363	0.370	-1.9	75	0.00
8 T	Diethyl Ether	0.224	0.238	-6.2	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.458	0.448	2.2	76	0.00
10 T	Methyl Iodide	0.750	0.779	-3.9	80	0.00
11 T	Tert butyl alcohol	0.036	0.032	11.1	74	0.01
12 CM	1,1-Dichloroethene	0.440	0.430	2.3#	76	0.00
13 T	Acrolein	0.032	0.027	15.6	74	0.00
14 T	Allyl chloride	0.751	0.683	9.1	73	0.00
15 T	Acrylonitrile	0.097	0.095	2.1	78	0.00
16 T	Acetone	0.097	0.078	19.6	63	0.00
17 T	Carbon Disulfide	1.331	1.312	1.4	75	0.00
18 T	Methyl Acetate	0.245	0.247	-0.8	78	0.00
19 T	Methyl tert-butyl Ether	0.685	0.697	-1.8	79	0.00
20 T	Methylene Chloride	0.486	0.459	5.6	79	0.00
21 T	trans-1,2-Dichloroethene	0.480	0.484	-0.8	77	0.00
22 T	Diisopropyl ether	1.403	1.352	3.6	76	0.00
23 T	Vinyl Acetate	0.864	0.818	5.3	74	0.00
24 P	1,1-Dichloroethane	0.859	0.838	2.4	76	0.00
25 T	2-Butanone	0.138	0.124	10.1	71	0.00
26 T	2,2-Dichloropropane	0.604	0.543	10.1	71	0.00
27 T	cis-1,2-Dichloroethene	0.543	0.545	-0.4	77	0.00
28 T	Bromochloromethane	0.317	0.301	5.0	79	0.00
29 T	Tetrahydrofuran	0.087	0.082	5.7	75	0.00
30 C	Chloroform	0.882	0.881	0.1#	78	0.00
31 T	Cyclohexane	0.850	0.755	11.2	72	0.00
32 T	1,1,1-Trichloroethane	0.785	0.773	1.5	77	0.00
33 S	1,2-Dichloroethane-d4	0.485	0.471	2.9	76	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.319	0.317	0.6	77	0.00
36 T	1,1-Dichloropropene	0.479	0.467	2.5	75	0.00
37 T	Ethyl Acetate	0.203	0.195	3.9	77	0.00
38 T	Carbon Tetrachloride	0.515	0.508	1.4	75	0.00
39 T	Methylcyclohexane	0.632	0.613	3.0	74	0.00
40 TM	Benzene	1.311	1.315	-0.3	77	0.00
41 T	Methacrylonitrile	0.127	0.128	-0.8	83	0.00
42 TM	1,2-Dichloroethane	0.396	0.389	1.8	77	0.00
43 T	Isopropyl Acetate	0.424	0.415	2.1	76	0.00
44 TM	Trichloroethene	0.380	0.390	-2.6	77	0.00
45 C	1,2-Dichloropropane	0.328	0.324	1.2#	77	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW100818\
 Data File : VW005908.D
 Acq On : 08 Oct 2018 21:10
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 6.68G/5ML/MSVOA_W/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 09 07:33:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 02:14:04 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.180	0.182	-1.1	79	0.00
47 T	Bromodichloromethane	0.454	0.450	0.9	78	0.00
48 T	Methyl methacrylate	0.202	0.197	2.5	76	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	82	0.00
50 S	Toluene-d8	1.228	1.223	0.4	77	0.00
51 T	4-Methyl-2-Pentanone	0.206	0.203	1.5	77	0.00
52 CM	Toluene	0.861	0.861	0.0	77	0.00
53 T	t-1,3-Dichloropropene	0.474	0.460	3.0	75	0.00
54 T	cis-1,3-Dichloropropene	0.533	0.517	3.0	75	0.00
55 T	1,1,2-Trichloroethane	0.250	0.255	-2.0	80	0.00
56 T	Ethyl methacrylate	0.337	0.335	0.6	78	0.00
57 T	1,3-Dichloropropane	0.420	0.422	-0.5	79	0.00
58 T	2-Chloroethyl Vinyl ether	0.137	0.129	5.8	77	0.00
59 T	2-Hexanone	0.148	0.141	4.7	73	0.00
60 T	Dibromochloromethane	0.322	0.328	-1.9	79	0.00
61 T	1,2-Dibromoethane	0.250	0.256	-2.4	80	0.00
62 S	4-Bromofluorobenzene	0.467	0.461	1.3	76	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	78	0.00
64 T	Tetrachloroethene	0.372	0.388	-4.3	81	0.00
65 PM	Chlorobenzene	1.071	1.072	-0.1	78	0.00
66 T	1,1,1,2-Tetrachloroethane	0.393	0.400	-1.8	79	0.00
67 C	Ethyl Benzene	1.892	1.884	0.4#	76	0.00
68 T	m/p-Xylenes	0.739	0.739	0.0	77	0.00
69 T	o-Xylene	0.708	0.717	-1.3	78	0.00
70 T	Styrene	1.174	1.183	-0.8	78	0.00
71 P	Bromoform	0.234	0.239	-2.1	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.652	3.501	4.1	77	0.00
74 T	N-amyl acetate	0.867	0.811	6.5	77	0.00
75 P	1,1,2,2-Tetrachloroethane	0.599	0.583	2.7	79	0.00
76 T	1,2,3-Trichloropropane	0.447	0.455	-1.8	82	0.00
77 T	Bromobenzene	0.865	0.866	-0.1	80	0.00
78 T	n-propylbenzene	4.296	4.103	4.5	76	0.00
79 T	2-Chlorotoluene	2.425	2.328	4.0	77	0.00
80 T	1,3,5-Trimethylbenzene	3.083	2.971	3.6	77	0.00
81 T	trans-1,4-Dichloro-2-butene	0.212	0.190	10.4	73	0.00
82 T	4-Chlorotoluene	2.551	2.480	2.8	77	0.00
83 T	tert-Butylbenzene	2.742	2.658	3.1	77	0.00
84 T	1,2,4-Trimethylbenzene	3.110	3.066	1.4	78	0.00
85 T	sec-Butylbenzene	3.849	3.755	2.4	77	0.00
86 T	p-Isopropyltoluene	3.461	3.397	1.8	77	0.00
87 T	1,3-Dichlorobenzene	1.719	1.718	0.1	79	0.00
88 T	1,4-Dichlorobenzene	1.708	1.688	1.2	79	0.00
89 T	n-Butylbenzene	3.263	3.135	3.9	76	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW100818\
Data File : VW005908.D
Acq On : 08 Oct 2018 21:10
Operator : SY/AP
Sample : VSTDCCC050
Misc : 6.68G/5ML/MSVOA_W/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Oct 09 07:33:52 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 02:14:04 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.651	0.627	3.7	76	0.00
91 T	1,2-Dichlorobenzene	1.526	1.542	-1.0	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.112	0.105	6.3	77	0.00
93 T	1,2,4-Trichlorobenzene	1.183	1.181	0.2	79	0.00
94 T	Hexachlorobutadiene	0.739	0.730	1.2	77	0.00
95 T	Naphthalene	2.039	2.083	-2.2	81	0.00
96 T	1,2,3-Trichlorobenzene	1.024	1.033	-0.9	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5