

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100819\
 Data File : VW013491.D
 Acq On : 08 Oct 2019 17:12
 Operator : SY/VA
 Sample : K5213-08RE
 Misc : 5.09G/5ML/MSVOA W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S-4(5-10)RE

Manual Integrations
 APPROVED

MMDadoda
 10/9/2019 4:10:18 PM

Quant Time: Oct 09 06:35:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 05:54:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	17287	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	23357	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	16703	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	5326	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	6830	51.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.28%	
35) Dibromofluoromethane	7.88	113	7279	55.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.60%	
50) Toluene-d8	10.33	98	25792	47.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.10%	
62) 4-Bromofluorobenzene	12.62	95	6084	33.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	66.08%	
Target Compounds						
16) Acetone	4.12	43	2273m	73.133	ug/l	Qvalue
17) Carbon Disulfide	4.40	76	3842	9.123	ug/l	96
20) Methylene Chloride	4.92	84	1895m	9.323	ug/l	
27) cis-1,2-Dichloroethene	7.17	96	10430	61.827	ug/l	95
44) Trichloroethene	9.11	130	1801	10.704	ug/l	40
64) Tetrachloroethene	10.86	164	6028	49.035	ug/l	92
85) sec-Butylbenzene	13.38	105	3603	9.263	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW100819\
Data File : VW013491.D
Acq On : 08 Oct 2019 17:12
Operator : SY/VA
Sample : K5213-08RE
Misc : 5.09G/5ML/MSVOA W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S-4(5-10)RE

Manual Integrations
APPROVED

MMDadoda
10/9/2019 4:10:18 PM

Quant Time: Oct 09 06:35:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100819S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 05:54:09 2019
Response via : Initial Calibration

