

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW100819\
 Data File : VW013493.D
 Acq On : 08 Oct 2019 18:29
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 09 06:40:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 05:54:09 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	0.166	0.183	-10.2	117	0.00
3 P	Chloromethane	0.291	0.292	-0.3	111	0.00
4 C	Vinyl Chloride	0.429	0.441	-2.8#	109	0.00
5 T	Bromomethane	0.256	0.280	-9.4	116	0.00
6 T	Chloroethane	0.241	0.260	-7.9	114	0.00
7 T	Trichlorofluoromethane	0.203	0.205	-1.0	98	0.00
8 T	Diethyl Ether	0.210	0.243	-15.7	126	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.419	0.438	-4.5	111	0.00
10 T	Methyl Iodide	0.663	0.728	-9.8	116	0.00
11 T	Tert butyl alcohol	0.033	0.032	3.0	119	0.00
12 CM	1,1-Dichloroethene	0.434	0.457	-5.3#	113	0.00
13 T	Acrolein	0.031	0.034	-9.7	103	0.00
14 T	Allyl chloride	0.667	0.725	-8.7	116	0.00
15 T	Acrylonitrile	0.098	0.110	-12.2	126	0.00
16 T	Acetone	0.090	0.097	-7.8	115	0.00
17 T	Carbon Disulfide	1.218	1.286	-5.6	112	0.00
18 T	Methyl Acetate	0.258	0.267	-3.5	125	0.00
19 T	Methyl tert-butyl Ether	0.656	0.763	-16.3	125	0.00
20 T	Methylene Chloride	0.495	0.497	-0.4	121	0.00
21 T	trans-1,2-Dichloroethene	0.461	0.497	-7.8	115	0.00
22 T	Diisopropyl ether	1.249	1.432	-14.7	122	0.00
23 T	Vinyl Acetate	0.750	0.874	-16.5	125	0.00
24 P	1,1-Dichloroethane	0.764	0.844	-10.5	119	0.00
25 T	2-Butanone	0.132	0.140	-6.1	118	0.00
26 T	2,2-Dichloropropane	0.492	0.493	-0.2	110	0.00
27 T	cis-1,2-Dichloroethene	0.488	0.541	-10.9	119	0.00
28 T	Bromochloromethane	0.272	0.318	-16.9	116	0.00
29 T	Tetrahydrofuran	0.081	0.091	-12.3	124	0.00
30 C	Chloroform	0.739	0.817	-10.6#	119	0.00
31 T	Cyclohexane	0.808	0.816	-1.0	113	0.00
32 T	1,1,1-Trichloroethane	0.600	0.655	-9.2	116	0.00
33 S	1,2-Dichloroethane-d4	0.386	0.423	-9.6	120	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	109	0.00
35 S	Dibromofluoromethane	0.279	0.299	-7.2	116	0.00
36 T	1,1-Dichloropropene	0.452	0.470	-4.0	114	0.00
37 T	Ethyl Acetate	0.205	0.222	-8.3	126	0.00
38 T	Carbon Tetrachloride	0.401	0.419	-4.5	111	0.00
39 T	Methylcyclohexane	0.586	0.622	-6.1	114	0.00
40 TM	Benzene	1.263	1.359	-7.6	118	0.00
41 T	Methacrylonitrile	0.117	0.130	-11.1	128	0.00
42 TM	1,2-Dichloroethane	0.319	0.359	-12.5	124	0.00
43 T	Isopropyl Acetate	0.371	0.417	-12.4	127	0.00
44 TM	Trichloroethene	0.360	0.389	-8.1	117	0.00
45 C	1,2-Dichloropropane	0.303	0.341	-12.5#	123	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW100819\
 Data File : VW013493.D
 Acq On : 08 Oct 2019 18:29
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 09 06:40:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 05:54:09 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.151	0.169	-11.9	125	0.00
47 T	Bromodichloromethane	0.375	0.429	-14.4	124	0.00
48 T	Methyl methacrylate	0.183	0.194	-6.0	120	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	110	0.00
50 S	Toluene-d8	1.174	1.238	-5.5	115	0.00
51 T	4-Methyl-2-Pentanone	0.195	0.214	-9.7	126	0.00
52 CM	Toluene	0.813	0.892	-9.7#	118	0.00
53 T	t-1,3-Dichloropropene	0.390	0.454	-16.4	127	0.00
54 T	cis-1,3-Dichloropropene	0.472	0.538	-14.0	124	0.00
55 T	1,1,2-Trichloroethane	0.227	0.255	-12.3	127	0.00
56 T	Ethyl methacrylate	0.305	0.355	-16.4	126	0.00
57 T	1,3-Dichloropropane	0.389	0.442	-13.6	127	0.00
58 T	2-Chloroethyl Vinyl ether	0.144	0.161	-11.8	118	0.00
59 T	2-Hexanone	0.134	0.149	-11.2	120	0.00
60 T	Dibromochloromethane	0.261	0.304	-16.5	125	0.00
61 T	1,2-Dibromoethane	0.216	0.246	-13.9	129	0.00
62 S	4-Bromofluorobenzene	0.394	0.436	-10.7	121	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	110	0.00
64 T	Tetrachloroethene	0.368	0.388	-5.4	117	0.00
65 PM	Chlorobenzene	0.994	1.074	-8.0	119	0.00
66 T	1,1,1,2-Tetrachloroethane	0.345	0.386	-11.9	123	0.00
67 C	Ethyl Benzene	1.816	1.944	-7.0#	118	0.00
68 T	m/p-Xylenes	0.691	0.754	-9.1	119	0.00
69 T	o-Xylene	0.638	0.711	-11.4	123	0.00
70 T	Styrene	1.096	1.233	-12.5	122	0.00
71 P	Bromoform	0.189	0.217	-14.8	127	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
73 T	Isopropylbenzene	3.582	3.810	-6.4	119	0.00
74 T	N-amyl acetate	0.783	0.874	-11.6	127	0.00
75 P	1,1,2,2-Tetrachloroethane	0.610	0.656	-7.5	126	0.00
76 T	1,2,3-Trichloropropane	0.416	0.417	-0.2	125	0.00
77 T	Bromobenzene	0.849	0.933	-9.9	126	0.00
78 T	n-propylbenzene	4.142	4.437	-7.1	120	0.00
79 T	2-Chlorotoluene	2.320	2.481	-6.9	121	0.00
80 T	1,3,5-Trimethylbenzene	2.976	3.211	-7.9	121	0.00
81 T	trans-1,4-Dichloro-2-butene	0.204	0.224	-9.8	126	0.00
82 T	4-Chlorotoluene	2.408	2.582	-7.2	121	0.00
83 T	tert-Butylbenzene	2.643	2.828	-7.0	121	0.00
84 T	1,2,4-Trimethylbenzene	2.943	3.166	-7.6	121	0.00
85 T	sec-Butylbenzene	3.652	3.853	-5.5	118	0.00
86 T	p-Isopropyltoluene	3.355	3.618	-7.8	120	0.00
87 T	1,3-Dichlorobenzene	1.610	1.786	-10.9	125	0.00
88 T	1,4-Dichlorobenzene	1.577	1.744	-10.6	126	0.00
89 T	n-Butylbenzene	3.031	3.283	-8.3	119	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW100819\
Data File : VW013493.D
Acq On : 08 Oct 2019 18:29
Operator : SY/VA
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Oct 09 06:40:40 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100819S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 05:54:09 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.574	0.616	-7.3	122	0.00
91 T	1,2-Dichlorobenzene	1.421	1.572	-10.6	126	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.097	0.105	-8.2	130	0.00
93 T	1,2,4-Trichlorobenzene	1.004	1.180	-17.5	131	0.00
94 T	Hexachlorobutadiene	0.667	0.746	-11.8	127	0.00
95 T	Naphthalene	1.738	2.106	-21.2#	137	0.00
96 T	1,2,3-Trichlorobenzene	0.878	1.027	-17.0	133	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6