

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100821\
 Data File : VW020416.D
 Acq On : 08 Oct 2021 18:27
 Operator : SY/VA
 Sample : M4103-08RE
 Misc : 4.72g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AC2RE

Quant Time: Oct 09 04:20:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 09 04:18:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	148010	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	103270	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	23167	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	55711	25.868	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.480%
7) Chloroethane-d5	2.879	69	31734	27.894	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	111.560%
11) 1,1-Dichloroethene-d2	4.013	63	71108	16.113	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.440%
21) 2-Butanone-d5	7.086	46	38626	58.931	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	117.860%
24) Chloroform-d	7.653	84	124866	26.466	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	105.880%
26) 1,2-Dichloroethane-d4	8.305	65	65357	24.255	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.000%
32) Benzene-d6	8.274	84	233967	34.443	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	137.760%#
36) 1,2-Dichloropropane-d6	9.274	67	84437	38.139	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	152.560%#
41) Toluene-d8	10.323	98	169655	27.835	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	111.360%
43) trans-1,3-Dichloroprop...	10.579	79	16134	17.933	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	71.720%
47) 2-Hexanone-d5	10.926	63	21355	62.703	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	125.400%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	56428	33.719	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	134.880%#
66) 1,2-Dichlorobenzene-d4	13.853	152	33984	37.500	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	150.000%#
Target Compounds						
					Qvalue	
13) Acetone	4.141	43	14531	24.147	ug/L	95
16) Methylene chloride	4.915	84	17392	6.045	ug/L	98
22) 2-Butanone	7.189	43	2679	3.557	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100821\
Data File : VW020416.D
Acq On : 08 Oct 2021 18:27
Operator : SY/VA
Sample : M4103-08RE
Misc : 4.72g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AC2RE

Quant Time: Oct 09 04:20:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 09 04:18:42 2021
Response via : Initial Calibration

