

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100821\
 Data File : VW020406.D
 Acq On : 08 Oct 2021 14:24
 Operator : SY/VA
 Sample : VW1008SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK456

Quant Time: Oct 09 04:19:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 09 04:18:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.847	114	201095	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.633	117	189744	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	87005	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.348	65	69680	23.813	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.240%
7) Chloroethane-d5	2.885	69	38791	25.097	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.400%
11) 1,1-Dichloroethene-d2	4.019	63	103239	17.218	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.880%
21) 2-Butanone-d5	7.085	46	41765	46.899	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.800%
24) Chloroform-d	7.652	84	150986	23.554	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.200%
26) 1,2-Dichloroethane-d4	8.305	65	88883	24.278	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.120%
32) Benzene-d6	8.280	84	299484	23.995	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.000%
36) 1,2-Dichloropropane-d6	9.274	67	97359	23.934	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.720%
41) Toluene-d8	10.323	98	274738	24.533	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.120%
43) trans-1,3-Dichloroprop...	10.579	79	38876	23.518	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.080%
47) 2-Hexanone-d5	10.926	63	27980	44.714	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.420%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	70004	22.767	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.080%
66) 1,2-Dichlorobenzene-d4	13.853	152	90609	26.622	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	106.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100821\
Data File : VW020406.D
Acq On : 08 Oct 2021 14:24
Operator : SY/VA
Sample : VW1008SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK456

Quant Time: Oct 09 04:19:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 09 04:18:42 2021
Response via : Initial Calibration

