

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100919\
 Data File : VW013511.D
 Acq On : 09 Oct 2019 18:18
 Operator : SY/VA
 Sample : K5292-04
 Misc : 4.98G/5ML/MSVOA W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 20190857-AMENDED-COMP-D

Quant Time: Oct 10 03:37:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 05:54:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	325970	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	486059	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	416606	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	191331	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	137851	54.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.48%	
35) Dibromofluoromethane	7.88	113	124773	45.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.92%	
50) Toluene-d8	10.32	98	561154	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
62) 4-Bromofluorobenzene	12.62	95	183819	47.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.92%	
Target Compounds						
16) Acetone	4.14	43	25136	42.890	ug/l	93
17) Carbon Disulfide	4.38	76	99978	12.590	ug/l	99
18) Methyl Acetate	4.68	43	2494	1.486	ug/l	93
25) 2-Butanone	7.18	43	6891	7.978	ug/l	91
95) Naphthalene	15.36	128	80497	12.107	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW100919\
Data File : VW013511.D
Acq On : 09 Oct 2019 18:18
Operator : SY/VA
Sample : K5292-04
Misc : 4.98G/5ML/MSVOA W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
20190857-AMENDED-COMP-D

Quant Time: Oct 10 03:37:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100819S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 05:54:09 2019
Response via : Initial Calibration

