

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW101018\
 Data File : VW005966.D
 Acq On : 10 Oct 2018 03:58
 Operator : SY/AP
 Sample : J5326-21
 Misc : 6.41G/5ML/MSVOA W/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 FL-PIT-6

Quant Time: Oct 10 05:53:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 02:14:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	248909	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	387460	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	345981	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	159073	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	120678	49.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.88%	
35) Dibromofluoromethane	7.88	113	122865	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
50) Toluene-d8	10.33	98	463296	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
62) 4-Bromofluorobenzene	12.62	95	151691	41.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.78%	
Target Compounds						
16) Acetone	4.14	43	25768	53.389	ug/l	97
20) Methylene Chloride	4.92	84	11135	4.605	ug/l	91
52) Toluene	10.39	92	18957	2.840	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW101018\
Data File : VW005966.D
Acq On : 10 Oct 2018 03:58
Operator : SY/AP
Sample : J5326-21
Misc : 6.41G/5ML/MSVOA W/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
FL-PIT-6

Quant Time: Oct 10 05:53:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 02:14:04 2018
Response via : Initial Calibration

