

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW101018\
 Data File : VW005971.D
 Acq On : 10 Oct 2018 06:07
 Operator : SY/AP
 Sample : J5326-39RE
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 FL-PH-3-STONERE

Quant Time: Oct 10 08:19:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 02:14:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	253403	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	386077	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	335205	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	155029	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	125369	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
35) Dibromofluoromethane	7.88	113	125375	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
50) Toluene-d8	10.33	98	446644	47.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.20%	
62) 4-Bromofluorobenzene	12.62	95	155027	42.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.94%	
Target Compounds						
20) Methylene Chloride	4.92	84	4850	1.970	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW101018\
Data File : VW005971.D
Acq On : 10 Oct 2018 06:07
Operator : SY/AP
Sample : J5326-39RE
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
FL-PH-3-STONERE

Quant Time: Oct 10 08:19:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 02:14:04 2018
Response via : Initial Calibration

