

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW101018\
 Data File : VW005993.D
 Acq On : 10 Oct 2018 17:21
 Operator : SY/AP
 Sample : J5306-08
 Misc : 6.90G/5ML/MSVOA W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S-20-V

Quant Time: Oct 11 05:19:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 02:14:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	185041	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	283638	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	254471	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	124289	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	103545	57.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.28%	
35) Dibromofluoromethane	7.88	113	99410	54.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.80%	
50) Toluene-d8	10.32	98	356380	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
62) 4-Bromofluorobenzene	12.62	95	126009	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
Target Compounds						
20) Methylene Chloride	4.92	84	66582	37.038	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW101018\
Data File : VW005993.D
Acq On : 10 Oct 2018 17:21
Operator : SY/AP
Sample : J5306-08
Misc : 6.90G/5ML/MSVOA W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S-20-V

Quant Time: Oct 11 05:19:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 02:14:04 2018
Response via : Initial Calibration

