

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW101018\
 Data File : VW005954.D
 Acq On : 09 Oct 2018 22:47
 Operator : SY/AP
 Sample : J5201-01
 Misc : 7.08G/5ML/MSVOA W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EO-02-100818-A

Quant Time: Oct 10 01:29:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 02:14:04 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	237131	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	363579	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	305271	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	129310	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	135138	58.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.40%	
35) Dibromofluoromethane	7.88	113	129257	55.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.36%	
50) Toluene-d8	10.33	98	448324	50.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.40%	
62) 4-Bromofluorobenzene	12.62	95	145449	42.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.62%	
Target Compounds						
20) Methylene Chloride	4.91	84	10089	4.379	ug/l	Qvalue 92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW101018\
Data File : VW005954.D
Acq On : 09 Oct 2018 22:47
Operator : SY/AP
Sample : J5201-01
Misc : 7.08G/5ML/MSVOA W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EO-02-100818-A

Quant Time: Oct 10 01:29:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 02:14:04 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 992

Delta R.T. -0.00 min

Lab File: VW005954.D

Acq: 09 Oct 2018 22:47

Instrument :

MSVOA_W

ClientSampleId :


EO-02-100818-A

Tot Ion:168 Resp: 237131

Ion Ratio Lower Upper

168 100

99 51.0 40.9 61.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW

7.95

120000

100000

80000



60000

40000

20000

0

Time--> 7.80 7.90 8.00 8.10



#20

Methylene Chloride

Concen: 4.379 ug/l

RT: 4.91 min Scan# 493

Delta R.T. -0.00 min

Lab File: VW005954.D

Acq: 09 Oct 2018 22:47

Tgt Ion: 84 Resp: 10089


Ion Ratio Lower Upper

84 100

49 110.3 99.9 149.9

51 35.5 31.1 46.7

86 63.7 51.8 77.6



Abundance Ion 83.90 (83.60 to 84.60): VW

Ion 48.90 (48.60 to 49.60): VW

Ion 50.90 (50.60 to 51.60): VW

Ion 85.90 (85.60 to 86.60): VW

4000


3000


2000

1000

0

Time--> 4.80 4.90 5.00 5.10









#33
 1,2-Dichloroethane-d4
 Concen: 58.704 ug/l
 RT: 8.30 min Scan# 1050
 Delta R.T. -0.01 min
 Lab File: VW005954.D
 Acq: 09 Oct 2018 22:47




Instrument :
 MSVOA_W
 ClientSampleId :
 EO-02-100818-A


Tot Ion: 65 Resp: 135138
 Ion Ratio Lower Upper
 65 100
 67 52.5 0.0 104.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.85 min Scan# 1139
 Delta R.T. 0.01 min
 Lab File: VW005954.D
 Acq: 09 Oct 2018 22:47

Tgt Ion: 114 Resp: 363579
 Ion Ratio Lower Upper
 114 100
 63 18.2 0.0 39.2
 88 14.0 0.0 30.8









#35
 Dibromofluoromethane
 Concen: 55.683 ug/l
 RT: 7.88 min Scan# 981
 Delta R.T. -0.00 min
 Lab File: VW005954.D
 Acq: 09 Oct 2018 22:47




Instrument :
 MSVOA_W
 ClientSampleId :
 EO-02-100818-A


Tot Ion	Ratio	Lower	Upper
113	100		
111	102.4	82.0	123.0
192	22.6	18.3	27.5



#50
 Toluene-d8
 Concen: 50.199 ug/l
 RT: 10.33 min Scan# 1382
 Delta R.T. -0.00 min
 Lab File: VW005954.D
 Acq: 09 Oct 2018 22:47

Tgt Ion	Ratio	Lower	Upper
98	100		
100	64.4	51.9	77.9









#62
4-Bromofluorobenzene
Concen: 42.805 ug/l
RT: 12.62 min Scan# 1758
Delta R.T. -0.00 min
Lab File: VW005954.D
Acq: 09 Oct 2018 22:47




Instrument :
MSVOA_W
ClientSampleId :
EO-02-100818-A


Tot Ion: 95 Resp: 145449
Ion Ratio Lower Upper
95 100
174 93.0 0.0 176.4
176 90.3 0.0 173.0



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.63 min Scan# 1596
Delta R.T. -0.00 min
Lab File: VW005954.D
Acq: 09 Oct 2018 22:47

Tgt Ion: 117 Resp: 305271
Ion Ratio Lower Upper
117 100
82 52.4 43.5 65.3
119 32.8 25.4 38.2





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.56 min Scan# 1912

Delta R.T. -0.01 min

Lab File: VW005954.D

Acq: 09 Oct 2018 22:47

Instrument :

MSVOA_W

ClientSampleId :

EO-02-100818-A



Tot Ion:152 Resp: 129310

Ion Ratio Lower Upper

152 100

115 55.7 28.3 85.0

150 156.2 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

