

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW101018\
 Data File : VW005984.D
 Acq On : 10 Oct 2018 13:28
 Operator : SY/AP
 Sample : J5255-01
 Misc : 6.38G/5ML/MSVOA W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TRACK-D-1

Quant Time: Oct 10 17:36:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 02:14:04 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	228152	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	344353	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	298291	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	134110	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	114499	51.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.40%	
35) Dibromofluoromethane	7.88	113	110636	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
50) Toluene-d8	10.33	98	395576	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
62) 4-Bromofluorobenzene	12.62	95	129890	40.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.72%	
Target Compounds						
20) Methylene Chloride	4.92	84	17758	8.012	ug/l	Qvalue 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW101018\
Data File : VW005984.D
Acq On : 10 Oct 2018 13:28
Operator : SY/AP
Sample : J5255-01
Misc : 6.38G/5ML/MSVOA W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TRACK-D-1

Quant Time: Oct 10 17:36:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 02:14:04 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 992

Delta R.T. 0.00 min

Lab File: VW005984.D

Acq: 10 Oct 2018 13:28

Instrument :

MSVOA_W

ClientSampleId :


TRACK-D-1

Tot Ion:168 Resp: 228152

Ion Ratio Lower Upper

168 100

99 50.5 40.9 61.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW

7.95

120000

100000

80000



60000

40000

20000

0

Time-->



#20

Methylene Chloride

Concen: 8.012 ug/l

RT: 4.92 min Scan# 494

Delta R.T. 0.01 min

Lab File: VW005984.D

Acq: 10 Oct 2018 13:28

Tgt Ion: 84 Resp: 17758


Ion Ratio Lower Upper

84 100

49 119.1 99.9 149.9

51 39.8 31.1 46.7

86 64.9 51.8 77.6



Abundance Ion 83.90 (83.60 to 84.60): VW

Ion 48.90 (48.60 to 49.60): VW

Ion 50.90 (50.60 to 51.60): VW

Ion 85.90 (85.60 to 86.60): VW

4.92

8000


6000


4000

2000

0

Time-->









#33
 1,2-Dichloroethane-d4
 Concen: 51.696 ug/l
 RT: 8.31 min Scan# 1050
 Delta R.T. -0.01 min
 Lab File: VW005984.D
 Acq: 10 Oct 2018 13:28




Instrument :
 MSVOA_W
 ClientSampleId :
 TRACK-D-1


Tot Ion: 65 Resp: 114499
 Ion Ratio Lower Upper
 65 100
 67 51.8 0.0 104.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.85 min Scan# 1139
 Delta R.T. 0.01 min
 Lab File: VW005984.D
 Acq: 10 Oct 2018 13:28

Tgt Ion: 114 Resp: 344353
 Ion Ratio Lower Upper
 114 100
 63 18.1 0.0 39.2
 88 14.7 0.0 30.8









#35
 Dibromofluoromethane
 Concen: 50.322 ug/l
 RT: 7.88 min Scan# 981
 Delta R.T. 0.00 min
 Lab File: VW005984.D
 Acq: 10 Oct 2018 13:28




Instrument :
 MSVOA_W
 ClientSampleId :
 TRACK-D-1


Tot Ion	Ratio	Lower	Upper
113	100		
111	101.8	82.0	123.0
192	22.4	18.3	27.5



#50
 Toluene-d8
 Concen: 46.765 ug/l
 RT: 10.33 min Scan# 1382
 Delta R.T. 0.00 min
 Lab File: VW005984.D
 Acq: 10 Oct 2018 13:28

Tgt Ion	Ratio	Lower	Upper
98	100		
100	63.8	51.9	77.9





#62

4-Bromofluorobenzene

Concen: 40.361 ug/l

RT: 12.62 min Scan# 1758

Delta R.T. 0.00 min

Lab File: VW005984.D

Acq: 10 Oct 2018 13:28

Instrument :

MSVOA_W

ClientSampled :

TRACK-D-1





Tot Ion: 95 Resp: 129890

Ion Ratio Lower Upper

95 100

174 92.5 0.0 176.4

176 89.5 0.0 173.0



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.63 min Scan# 1596

Delta R.T. 0.00 min

Lab File: VW005984.D

Acq: 10 Oct 2018 13:28




Tgt Ion: 117 Resp: 298291


Ion Ratio Lower Upper

117 100

82 52.9 43.5 65.3

119 32.1 25.4 38.2





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.57 min Scan# 1913

Delta R.T. 0.00 min

Lab File: VW005984.D

Acq: 10 Oct 2018 13:28

Instrument :

MSVOA_W

ClientSampleId :

TRACK-D-1


Tot Ion:152 Resp: 134110

Ion Ratio Lower Upper

152 100

115 55.1 28.3 85.0

150 155.9 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

