


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW101018\
Data File : VW005947.D
Acq On : 09 Oct 2018 18:54
Operator : SY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
Title : SW846 8260
Last Update : Tue Oct 09 02:14:04 2018



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	25629	PASS
75	95	30	60	51.4	69845	PASS
95	95	100	100	100.0	135789	PASS
96	95	5	9	6.7	9126	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	91.2	123834	PASS
175	174	5	9	7.7	9533	PASS
176	174	95	101	96.5	119472	PASS
177	176	5	9	7.1	8491	PASS