

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101024\
 Data File : VW030522.D
 Acq On : 10 Oct 2024 21:05
 Operator : SY/MD
 Sample : P4346-20
 Misc : 3.20g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EOAY0

Quant Time: Oct 11 03:25:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 11 02:35:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	105564	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	115554	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	49391	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.356	65	33223	20.022	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.080%
7) Chloroethane-d5	2.893	69	29779	23.709	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.840%
11) 1,1-Dichloroethene-d2	4.015	65	12628	18.298	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.200%
21) 2-Butanone-d5	7.081	46	23518	78.758	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	157.520%#
24) Chloroform-d	7.648	84	70669	23.409	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.640%
26) 1,2-Dichloroethane-d4	8.307	65	50439	32.130	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	128.520%
32) Benzene-d6	8.276	84	124738	18.599	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	74.400%
36) 1,2-Dichloropropane-d6	9.276	67	43993	22.349	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.400%
41) Toluene-d8	10.324	98	89478	14.615	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	58.440%
43) trans-1,3-Dichloroprop...	10.581	79	14698	17.831	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	71.320%
47) 2-Hexanone-d5	10.928	63	14692	57.947	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	115.900%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	36818	25.468	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	101.880%
66) 1,2-Dichlorobenzene-d4	13.848	152	19594	11.725	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	46.880%#
Target Compounds					Qvalue	
13) Acetone	4.124	43	6443	14.738	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101024\
Data File : VW030522.D
Acq On : 10 Oct 2024 21:05
Operator : SY/MD
Sample : P4346-20
Misc : 3.20g/10mL/MSVOA_W/SOIL/B
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EOAY0

Quant Time: Oct 11 03:25:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 11 02:35:02 2024
Response via : Initial Calibration

