

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101024\
 Data File : VW030528.D
 Acq On : 10 Oct 2024 23:30
 Operator : SY/MD
 Sample : P4351-06
 Misc : 5.44g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EOLG8

Quant Time: Oct 11 03:27:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 11 02:35:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	354280	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	258619	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	68979	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	168301	30.222	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 120.880%	
7) Chloroethane-d5	2.899	69	138529	32.863	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 131.440%	
11) 1,1-Dichloroethene-d2	4.021	65	61552	26.575	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 106.280%	
21) 2-Butanone-d5	7.075	46	81359	81.183	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 162.360%#	
24) Chloroform-d	7.648	84	279990	27.635	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 110.560%	
26) 1,2-Dichloroethane-d4	8.301	65	166427	31.589	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 126.360%	
32) Benzene-d6	8.276	84	543103	36.182	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 144.720%#	
36) 1,2-Dichloropropane-d6	9.270	67	163161	37.035	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 148.120%#	
41) Toluene-d8	10.325	98	436524	31.857	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 127.440%	
43) trans-1,3-Dichloroprop...	10.575	79	58328	31.618	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 126.480%	
47) 2-Hexanone-d5	10.922	63	54480	96.010	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 192.020%#	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	119543	36.947	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 147.800%#	
66) 1,2-Dichlorobenzene-d4	13.849	152	97893	41.943	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 167.760%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101024\
Data File : VW030528.D
Acq On : 10 Oct 2024 23:30
Operator : SY/MD
Sample : P4351-06
Misc : 5.44g/10mL/MSVOA_W/SOIL/A
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EOLG8

Quant Time: Oct 11 03:27:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 11 02:35:02 2024
Response via : Initial Calibration

