

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101024\
 Data File : VW030503.D
 Acq On : 10 Oct 2024 11:11
 Operator : SY/MD
 Sample : P4346-05
 Misc : 2.61g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EOAW7

Quant Time: Oct 11 01:21:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 11 01:19:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	544510	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	525505	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	240797	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	209949	24.530	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.120%
7) Chloroethane-d5	2.893	69	171972	26.544	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.160%
11) 1,1-Dichloroethene-d2	4.021	65	82994	23.314	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	93.240%
21) 2-Butanone-d5	7.075	46	102706	66.680	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	133.360%
24) Chloroform-d	7.648	84	385527	24.758	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.040%
26) 1,2-Dichloroethane-d4	8.307	65	230750	28.497	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	114.000%
32) Benzene-d6	8.270	84	734199	24.072	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.280%
36) 1,2-Dichloropropane-d6	9.276	67	226408	25.291	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.160%
41) Toluene-d8	10.325	98	688248	24.719	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.880%
43) trans-1,3-Dichloroprop...	10.575	79	101244	27.009	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	108.040%
47) 2-Hexanone-d5	10.922	63	79312	68.786	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	137.580%#
56) 1,1,2,2-Tetrachloroeth...	12.690	84	193676	29.459	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	117.840%
66) 1,2-Dichlorobenzene-d4	13.848	152	219398	26.928	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	107.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101024\
Data File : VW030503.D
Acq On : 10 Oct 2024 11:11
Operator : SY/MD
Sample : P4346-05
Misc : 2.61g/10mL/MSVOA_W/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EOAW7

Quant Time: Oct 11 01:21:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 11 01:19:48 2024
Response via : Initial Calibration

