

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101222\
 Data File : VW024880.D
 Acq On : 12 Oct 2022 17:58
 Operator : SY/VA
 Sample : N5070-06
 Misc : 5.05g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 COBM2

Quant Time: Oct 12 23:22:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 12 23:18:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	519512	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	409976	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	144522	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	119565	19.508	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.040%
7) Chloroethane-d5	2.885	69	74617	20.793	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.160%
11) 1,1-Dichloroethene-d2	4.019	63	176076	13.213	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	52.840%
21) 2-Butanone-d5	7.086	46	72549	37.358	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.720%
24) Chloroform-d	7.647	84	355740	24.277	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.120%
26) 1,2-Dichloroethane-d4	8.305	65	196913	25.677	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	102.720%
32) Benzene-d6	8.275	84	625278	24.322	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.280%
36) 1,2-Dichloropropane-d6	9.274	67	228036	28.920	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	115.680%
41) Toluene-d8	10.323	98	490522	21.258	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.040%
43) trans-1,3-Dichloroprop...	10.579	79	34071	11.393	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	45.560%
47) 2-Hexanone-d5	10.927	63	25617	19.377	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	38.760%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	195552	31.800	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	127.200%#
66) 1,2-Dichlorobenzene-d4	13.853	152	126927	24.033	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.120%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101222\
Data File : VW024880.D
Acq On : 12 Oct 2022 17:58
Operator : SY/VA
Sample : N5070-06
Misc : 5.05g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BM2

Quant Time: Oct 12 23:22:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 12 23:18:51 2022
Response via : Initial Calibration

