

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101222\
 Data File : VW024876.D
 Acq On : 12 Oct 2022 16:22
 Operator : SY/VA
 Sample : N5070-02
 Misc : 6.44g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0BK5

Quant Time: Oct 12 23:21:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 12 23:18:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	496409	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	412348	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	152595	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	118448	20.225	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.920%
7) Chloroethane-d5	2.885	69	81174	23.673	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.680%
11) 1,1-Dichloroethene-d2	4.019	63	189510	14.883	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.520%
21) 2-Butanone-d5	7.080	46	97135	52.346	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.700%
24) Chloroform-d	7.647	84	353302	25.233	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.920%
26) 1,2-Dichloroethane-d4	8.305	65	197194	26.910	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	107.640%
32) Benzene-d6	8.275	84	643903	24.902	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.600%
36) 1,2-Dichloropropane-d6	9.274	67	227118	28.638	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	114.560%
41) Toluene-d8	10.323	98	505593	21.785	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.160%
43) trans-1,3-Dichloroprop...	10.579	79	63792	21.209	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.840%
47) 2-Hexanone-d5	10.927	63	64976	48.867	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.740%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	198029	32.018	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	128.080%#
66) 1,2-Dichlorobenzene-d4	13.847	152	140123	25.128	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101222\
Data File : VW024876.D
Acq On : 12 Oct 2022 16:22
Operator : SY/VA
Sample : N5070-02
Misc : 6.44g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BK5

Quant Time: Oct 12 23:21:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 12 23:18:51 2022
Response via : Initial Calibration

