

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW101318\
 Data File : VW006058.D
 Acq On : 12 Oct 2018 20:20
 Operator : VA/AP
 Sample : J5198-15RE
 Misc : 10.78G/5ML/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CL-02-101218-CRE

Quant Time: Oct 13 02:23:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 02:14:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	206266	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	298536	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	265089	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	120679	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	94073	46.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.96%	
35) Dibromofluoromethane	7.89	113	94077	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
50) Toluene-d8	10.33	98	336461	45.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.76%	
62) 4-Bromofluorobenzene	12.62	95	115517	41.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.80%	
Target Compounds						
16) Acetone	4.13	43	14199	35.50	ug/l	Qvalue 97
20) Methylene Chloride	4.92	84	32731	16.33	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW101318\
Data File : VW006058.D
Acq On : 12 Oct 2018 20:20
Operator : VA/AP
Sample : J5198-15RE
Misc : 10.78G/5ML/MSVOA_W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CL-02-101218-CRE

Quant Time: Oct 13 02:23:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 02:14:04 2018
Response via : Initial Calibration

