


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101318\
Data File : VW006040.D
Acq On : 12 Oct 2018 09:54
Operator : VA/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100818S.M
Title : SW846 8260
Last Update : Tue Oct 09 02:14:04 2018



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1747

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	27330	PASS
75	95	30	60	50.9	73250	PASS
95	95	100	100	100.0	143905	PASS
96	95	5	9	6.7	9699	PASS
173	174	0.00	2	0.4	584	PASS
174	95	50	100	90.3	129944	PASS
175	174	5	9	7.9	10233	PASS
176	174	95	101	97.5	126656	PASS
177	176	5	9	7.0	8847	PASS