

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW101421\
 Data File : VW022825.D
 Acq On : 15 Oct 2021 05:03
 Operator : SY/MD
 Sample : M4223-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFGC6

Quant Time: Oct 15 06:47:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 15 06:41:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	94243	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	94281	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	46271	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	35644	5.497	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.000%
7) Chloroethane-d5	1.568	69	32896	5.646	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.000%
11) 1,1-Dichloroethene-d2	2.111	63	41659	3.275	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.600%
20) 2-Butanone-d5	3.889	46	65811	38.709	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.420%
24) Chloroform-d	4.352	84	69643	5.212	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.200%
26) 1,2-Dichloroethane-d4	5.034	65	32594	5.026	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.600%
32) Benzene-d6	5.053	84	130749	4.659	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.200%
36) 1,2-Dichloropropane-d6	6.069	67	39371	4.737	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.800%
41) Toluene-d8	7.317	98	114304	4.711	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.200%
43) trans-1,3-Dichloroprop...	7.625	79	11751	4.262	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.200%
46) 2-Hexanone-d5	8.092	63	53709	48.335	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.680%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	26614	4.727	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	45914	5.352	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.000%
Target Compounds						
					Qvalue	
13) Acetone	2.182	43	30720	32.404	ug/L	# 51
42) Toluene	7.397	91	5154	0.200	ug/L	92
47) Tetrachloroethene	7.982	164	2435	0.436	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101421\
Data File : VV022825.D
Acq On : 15 Oct 2021 05:03
Operator : SY/MD
Sample : M4223-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFGC6

Quant Time: Oct 15 06:47:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 15 06:41:28 2021
Response via : Initial Calibration

