

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101421\
 Data File : VW020496.D
 Acq On : 14 Oct 2021 15:58
 Operator : SY/VA
 Sample : M4205-05
 Misc : 6.21g/10mL/MSVOA_W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0B01

Quant Time: Oct 14 23:37:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 14 23:35:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	47923	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	44482	25.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	16345	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	19038	27.30	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	109.20%
7) Chloroethane-d5	2.879	69	13448	36.51	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	146.04%
11) 1,1-Dichloroethene-d2	4.013	63	29223	20.45	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	81.80%
21) 2-Butanone-d5	7.086	46	9089	42.83	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.66%
24) Chloroform-d	7.653	84	40760	26.68	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	106.72%
26) 1,2-Dichloroethane-d4	8.311	65	22897	26.24	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.96%
32) Benzene-d6	8.281	84	72994	24.95	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.80%
36) 1,2-Dichloropropane-d6	9.275	67	24959	26.17	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.68%
41) Toluene-d8	10.329	98	57346	21.84	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.36%
43) trans-1,3-Dichloroprop...	10.579	79	7613	19.65	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.60%
47) 2-Hexanone-d5	10.927	63	4999	34.08	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	68.16%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	16816	23.33	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.32%
66) 1,2-Dichlorobenzene-d4	13.853	152	16096	25.17	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.68%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101421\
Data File : VW020496.D
Acq On : 14 Oct 2021 15:58
Operator : SY/VA
Sample : M4205-05
Misc : 6.21g/10mL/MSVOA_W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0B01

Quant Time: Oct 14 23:37:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 14 23:35:04 2021
Response via : Initial Calibration

