

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101421\
 Data File : VW020503.D
 Acq On : 14 Oct 2021 20:24
 Operator : SY/VA
 Sample : M4205-10
 Misc : 5.63g/10mL/MSVOA_W/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0B35

Quant Time: Oct 15 00:59:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 15 00:58:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	215053	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	191254	25.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	77034	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	66714	21.32	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.28%
7) Chloroethane-d5	2.879	69	37283	22.56	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.24%
11) 1,1-Dichloroethene-d2	4.013	63	96944	15.12	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.48%
21) 2-Butanone-d5	7.080	46	33925	35.62	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	71.24%
24) Chloroform-d	7.653	84	152569	22.26	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.04%
26) 1,2-Dichloroethane-d4	8.305	65	83160	21.24	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.96%
32) Benzene-d6	8.275	84	287646	22.86	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.44%
36) 1,2-Dichloropropane-d6	9.274	67	95862	23.38	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.52%
41) Toluene-d8	10.323	98	247216	21.90	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.60%
43) trans-1,3-Dichloroprop...	10.579	79	30215	18.13	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	72.52%
47) 2-Hexanone-d5	10.927	63	20428	32.39	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	64.78%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	61340	19.79	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	79.16%
66) 1,2-Dichlorobenzene-d4	13.853	152	70980	23.55	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	94.20%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101421\
Data File : VW020503.D
Acq On : 14 Oct 2021 20:24
Operator : SY/VA
Sample : M4205-10
Misc : 5.63g/10mL/MSVOA_W/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0B35

Quant Time: Oct 15 00:59:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 15 00:58:32 2021
Response via : Initial Calibration

