

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101421\
 Data File : VW020512.D
 Acq On : 15 Oct 2021 00:00
 Operator : SY/VA
 Sample : M4207-02
 Misc : 6.10g/10mL/MSVOA_W/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0B48

Quant Time: Oct 15 01:00:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 15 00:58:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	207789	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	192446	25.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	80305	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	65038	21.51	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.04%
7) Chloroethane-d5	2.879	69	37132	23.25	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.00%
11) 1,1-Dichloroethene-d2	4.019	63	96499	15.58	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.32%
21) 2-Butanone-d5	7.080	46	34916	37.95	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.90%
24) Chloroform-d	7.653	84	157776	23.82	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.28%
26) 1,2-Dichloroethane-d4	8.311	65	90556	23.94	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	95.76%
32) Benzene-d6	8.281	84	294175	23.24	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.96%
36) 1,2-Dichloropropane-d6	9.280	67	99835	24.20	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.80%
41) Toluene-d8	10.323	98	254197	22.38	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.52%
43) trans-1,3-Dichloroprop...	10.579	79	32570	19.43	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	77.72%
47) 2-Hexanone-d5	10.926	63	21226	33.44	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	66.88%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	70629	22.65	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	90.60%
66) 1,2-Dichlorobenzene-d4	13.853	152	76393	24.32	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.28%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101421\
Data File : VW020512.D
Acq On : 15 Oct 2021 00:00
Operator : SY/VA
Sample : M4207-02
Misc : 6.10g/10mL/MSVOA_W/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0B48

Quant Time: Oct 15 01:00:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 15 00:58:32 2021
Response via : Initial Calibration

