

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
 Data File : VW020527.D
 Acq On : 15 Oct 2021 12:16
 Operator : SY/VA
 Sample : VW1015SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK464

Quant Time: Oct 16 06:56:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 16 06:56:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	186931	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	178626	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	83757	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	58612	21.548	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.200%
7) Chloroethane-d5	2.879	69	33195	23.103	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.400%
11) 1,1-Dichloroethene-d2	4.013	63	89856	16.122	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.480%
21) 2-Butanone-d5	7.080	46	40766	49.246	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.500%
24) Chloroform-d	7.653	84	129081	21.663	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.640%
26) 1,2-Dichloroethane-d4	8.305	65	77484	22.768	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.080%
32) Benzene-d6	8.275	84	250006	21.278	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.120%
36) 1,2-Dichloropropane-d6	9.274	67	82182	21.461	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.840%
41) Toluene-d8	10.323	98	220739	20.938	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.760%
43) trans-1,3-Dichloroprop...	10.579	79	32560	20.924	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.680%
47) 2-Hexanone-d5	10.927	63	27673	46.976	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.960%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	64612	22.322	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.280%
66) 1,2-Dichlorobenzene-d4	13.853	152	75198	22.951	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
Data File : VW020527.D
Acq On : 15 Oct 2021 12:16
Operator : SY/VA
Sample : VW1015SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK464

Quant Time: Oct 16 06:56:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 16 06:56:08 2021
Response via : Initial Calibration

