

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
 Data File : VW020528.D
 Acq On : 15 Oct 2021 12:58
 Operator : SY/VA
 Sample : M4205-05
 Misc : 6.24g/10mL/MSVOA_W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0B01

Quant Time: Oct 16 06:56:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 16 06:56:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	258837	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	232837	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	97724	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	74413	19.757	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.040%
7) Chloroethane-d5	2.879	69	45521	22.881	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.520%
11) 1,1-Dichloroethene-d2	4.007	63	117329	15.203	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.800%
21) 2-Butanone-d5	7.086	46	44581	38.894	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.780%
24) Chloroform-d	7.647	84	181368	21.982	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.920%
26) 1,2-Dichloroethane-d4	8.305	65	100349	21.295	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.200%
32) Benzene-d6	8.275	84	343739	22.444	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.760%
36) 1,2-Dichloropropane-d6	9.275	67	113125	22.663	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.640%
41) Toluene-d8	10.323	98	297516	21.650	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.600%
43) trans-1,3-Dichloroprop...	10.579	79	41903	20.658	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.640%
47) 2-Hexanone-d5	10.927	63	30585	39.831	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.660%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	75342	19.968	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	79.880%
66) 1,2-Dichlorobenzene-d4	13.853	152	88027	23.027	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.120%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
Data File : VW020528.D
Acq On : 15 Oct 2021 12:58
Operator : SY/VA
Sample : M4205-05
Misc : 6.24g/10mL/MSVOA_W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0B01

Quant Time: Oct 16 06:56:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 16 06:56:08 2021
Response via : Initial Calibration

