

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
 Data File : VW020539.D
 Acq On : 15 Oct 2021 17:23
 Operator : SY/VA
 Sample : M4210-20
 Misc : 4.83g/10mL/MSVOA_W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JDGF2

Quant Time: Oct 16 06:57:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 16 06:56:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	257387	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	230232	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	86894	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	69767	18.628	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.520%
7) Chloroethane-d5	2.879	69	40345	20.393	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.560%
11) 1,1-Dichloroethene-d2	4.013	63	107599	14.021	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	56.080%
21) 2-Butanone-d5	7.086	46	42459	37.251	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.500%
24) Chloroform-d	7.647	84	175738	21.420	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.680%
26) 1,2-Dichloroethane-d4	8.305	65	102020	21.772	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.080%
32) Benzene-d6	8.275	84	331203	21.870	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.480%
36) 1,2-Dichloropropane-d6	9.274	67	111283	22.546	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.200%
41) Toluene-d8	10.323	98	291161	21.427	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.720%
43) trans-1,3-Dichloroprop...	10.579	79	38707	19.298	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	77.200%
47) 2-Hexanone-d5	10.927	63	23984	31.588	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	63.180%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	73533	19.710	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	78.840%
66) 1,2-Dichlorobenzene-d4	13.853	152	75304	22.154	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
Data File : VW020539.D
Acq On : 15 Oct 2021 17:23
Operator : SY/VA
Sample : M4210-20
Misc : 4.83g/10mL/MSVOA_W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JDGF2

Quant Time: Oct 16 06:57:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 16 06:56:08 2021
Response via : Initial Calibration

