

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
 Data File : VW020541.D
 Acq On : 15 Oct 2021 18:11
 Operator : SY/VA
 Sample : M4210-22
 Misc : 2.88g/10mL/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JDGG1

Quant Time: Oct 16 06:58:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 16 06:56:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	199117	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	163712	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	60165	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	61382	21.186	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.760%
7) Chloroethane-d5	2.879	69	37115	24.251	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.000%
11) 1,1-Dichloroethene-d2	4.013	63	88422	14.894	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.560%
21) 2-Butanone-d5	7.080	46	35678	40.462	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.920%
24) Chloroform-d	7.653	84	143842	22.663	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.640%
26) 1,2-Dichloroethane-d4	8.305	65	81538	22.493	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.960%
32) Benzene-d6	8.275	84	264291	24.543	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	98.160%
36) 1,2-Dichloropropane-d6	9.275	67	89726	25.565	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	102.280%
41) Toluene-d8	10.323	98	212971	22.042	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.160%
43) trans-1,3-Dichloroprop...	10.579	79	30388	21.307	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.240%
47) 2-Hexanone-d5	10.927	63	23114	42.811	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.620%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	54360	20.491	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.960%
66) 1,2-Dichlorobenzene-d4	13.853	152	48950	20.798	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	83.200%
Target Compounds					Qvalue	
13) Acetone	4.129	43	82284	101.642	ug/L	98
22) 2-Butanone	7.177	43	20542	20.271	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
Data File : VW020541.D
Acq On : 15 Oct 2021 18:11
Operator : SY/VA
Sample : M4210-22
Misc : 2.88g/10mL/MSVOA_W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JDGG1

Quant Time: Oct 16 06:58:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 16 06:56:08 2021
Response via : Initial Calibration

