

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
 Data File : VW020534.D
 Acq On : 15 Oct 2021 15:22
 Operator : SY/VA
 Sample : M4210-12
 Misc : 3.74g/10mL/MSVOA_W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JDGE0

Quant Time: Oct 16 06:57:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 16 06:56:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	238257	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	228708	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	105733	25.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	69615	20.080	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.320%
7) Chloroethane-d5	2.879	69	42316	23.107	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.440%
11) 1,1-Dichloroethene-d2	4.013	63	106963	15.057	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.240%
21) 2-Butanone-d5	7.080	46	44846	42.504	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.000%
24) Chloroform-d	7.653	84	163762	21.563	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.240%
26) 1,2-Dichloroethane-d4	8.305	65	94724	21.838	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.360%
32) Benzene-d6	8.275	84	313623	20.847	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	83.400%
36) 1,2-Dichloropropane-d6	9.274	67	105041	21.424	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.680%
41) Toluene-d8	10.323	98	277330	20.546	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	82.200%
43) trans-1,3-Dichloroprop...	10.579	79	39183	19.666	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.680%
47) 2-Hexanone-d5	10.927	63	29887	39.624	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.240%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	72779	19.637	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	78.560%
66) 1,2-Dichlorobenzene-d4	13.853	152	90886	21.974	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
Data File : VW020534.D
Acq On : 15 Oct 2021 15:22
Operator : SY/VA
Sample : M4210-12
Misc : 3.74g/10mL/MSVOA_W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JDGE0

Quant Time: Oct 16 06:57:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 16 06:56:08 2021
Response via : Initial Calibration

