

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
 Data File : VW020535.D
 Acq On : 15 Oct 2021 15:47
 Operator : SY/VA
 Sample : M4210-16
 Misc : 5.29g/10mL/MSVOA_W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JDGE5

Quant Time: Oct 16 06:57:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 16 06:56:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	268663	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	247993	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	116003	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	69809	17.857	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.440%
7) Chloroethane-d5	2.879	69	45116	21.848	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.400%
11) 1,1-Dichloroethene-d2	4.013	63	107208	13.383	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	53.520%
21) 2-Butanone-d5	7.086	46	44606	37.492	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.980%
24) Chloroform-d	7.653	84	167911	19.607	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.440%
26) 1,2-Dichloroethane-d4	8.305	65	97800	19.995	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	80.000%
32) Benzene-d6	8.275	84	320098	19.623	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	78.480%
36) 1,2-Dichloropropane-d6	9.274	67	107590	20.237	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	80.960%
41) Toluene-d8	10.323	98	280557	19.168	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	76.680%
43) trans-1,3-Dichloroprop...	10.579	79	39777	18.411	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	73.640%
47) 2-Hexanone-d5	10.927	63	30338	37.095	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.180%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	75991	18.910	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	75.640%
66) 1,2-Dichlorobenzene-d4	13.853	152	89334	19.687	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
Data File : VW020535.D
Acq On : 15 Oct 2021 15:47
Operator : SY/VA
Sample : M4210-16
Misc : 5.29g/10mL/MSVOA_W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JDGE5

Quant Time: Oct 16 06:57:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 16 06:56:08 2021
Response via : Initial Calibration

