

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
 Data File : VW020540.D
 Acq On : 15 Oct 2021 17:47
 Operator : SY/VA
 Sample : M4210-21
 Misc : 5.03g/10mL/MSVOA_W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JDGF3

Quant Time: Oct 16 06:57:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 16 06:56:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	224894	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	193979	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	75232	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	69270	21.168	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.680%
7) Chloroethane-d5	2.879	69	44536	25.764	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.040%
11) 1,1-Dichloroethene-d2	4.013	63	102267	15.251	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	61.000%
21) 2-Butanone-d5	7.086	46	40797	40.964	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.920%
24) Chloroform-d	7.653	84	162711	22.697	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.800%
26) 1,2-Dichloroethane-d4	8.305	65	93298	22.787	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.160%
32) Benzene-d6	8.274	84	305568	23.948	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.800%
36) 1,2-Dichloropropane-d6	9.274	67	103235	24.825	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.280%
41) Toluene-d8	10.323	98	263048	22.976	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.920%
43) trans-1,3-Dichloroprop...	10.579	79	35870	21.226	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.920%
47) 2-Hexanone-d5	10.926	63	26401	41.269	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.540%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	68180	21.690	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	86.760%
66) 1,2-Dichlorobenzene-d4	13.853	152	68459	23.262	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.040%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101521\
Data File : VW020540.D
Acq On : 15 Oct 2021 17:47
Operator : SY/VA
Sample : M4210-21
Misc : 5.03g/10mL/MSVOA_W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JDGF3

Quant Time: Oct 16 06:57:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 16 06:56:08 2021
Response via : Initial Calibration

