

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW101618\
 Data File : VW006101.D
 Acq On : 17 Oct 2018 03:04
 Operator : VA/AP
 Sample : J5526-20
 Misc : 6.9G/5ML/MSVOA W/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SUP-4A-G-101518-5

Quant Time: Oct 17 07:43:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 16 05:36:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	283698	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	421021	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	370034	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	197326	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	109124	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
35) Dibromofluoromethane	7.88	113	109805	44.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.28%	
50) Toluene-d8	10.33	98	356810	37.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.40%	
62) 4-Bromofluorobenzene	12.62	95	150186	42.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.50%	
Target Compounds						
16) Acetone	4.14	43	9010	5.968	ug/l	98
20) Methylene Chloride	4.92	84	18343	5.936	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW101618\
Data File : VW006101.D
Acq On : 17 Oct 2018 03:04
Operator : VA/AP
Sample : J5526-20
Misc : 6.9G/5ML/MSVOA W/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SUP-4A-G-101518-5

Quant Time: Oct 17 07:43:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 16 05:36:33 2018
Response via : Initial Calibration

