

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW101618\
 Data File : VW006104.D
 Acq On : 17 Oct 2018 04:22
 Operator : VA/AP
 Sample : J5539-01
 Misc : 7.4G/5ML/MSVOA W/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-6-0.3-2FT

Quant Time: Oct 17 07:48:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 16 05:36:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	200283	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	301512	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	257312	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	119076	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	95692	57.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.26%	
35) Dibromofluoromethane	7.88	113	89555	50.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.54%	
50) Toluene-d8	10.33	98	322759	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
62) 4-Bromofluorobenzene	12.62	95	108069	42.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.90%	
Target Compounds						
16) Acetone	4.13	43	17002	42.278	ug/l	Qvalue 100
20) Methylene Chloride	4.92	84	12027	5.393	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW101618\
Data File : VW006104.D
Acq On : 17 Oct 2018 04:22
Operator : VA/AP
Sample : J5539-01
Misc : 7.4G/5ML/MSVOA W/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-6-0.3-2FT

Quant Time: Oct 17 07:48:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 16 05:36:33 2018
Response via : Initial Calibration

