

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW101618\
 Data File : VW006110.D
 Acq On : 17 Oct 2018 06:56
 Operator : VA/AP
 Sample : J5539-07
 Misc : 11.91G/5ML/MSVOA W/SOIL
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-10-5-7FT

Quant Time: Oct 17 12:25:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 16 05:36:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	19518	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	29734	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	25342	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	18018	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	9166	56.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.28%	
35) Dibromofluoromethane	7.89	113	8577	48.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.64%	
50) Toluene-d8	10.33	98	32017	47.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.80%	
62) 4-Bromofluorobenzene	12.62	95	19253	76.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	153.38%	
Target Compounds						
16) Acetone	4.12	43	4774	151.422	ug/l	98
20) Methylene Chloride	4.93	84	1777	9.045	ug/l	86
73) Isopropylbenzene	12.47	105	3341	2.625	ug/l	82
78) n-propylbenzene	12.81	91	8285	5.594	ug/l	89
83) tert-Butylbenzene	13.21	119	6026	6.276	ug/l	79
85) sec-Butylbenzene	13.39	105	24898	18.356	ug/l #	54

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW101618\
Data File : VW006110.D
Acq On : 17 Oct 2018 06:56
Operator : VA/AP
Sample : J5539-07
Misc : 11.91G/5ML/MSVOA W/SOIL
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-10-5-7FT

Quant Time: Oct 17 12:25:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 16 05:36:33 2018
Response via : Initial Calibration

