

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW101618\
 Data File : VW006118.D
 Acq On : 17 Oct 2018 10:23
 Operator : VA/AP
 Sample : J5200-11
 Misc : 6.70G/5ML/MSVOA W/SOIL
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 OR-01-101618-B

Quant Time: Oct 17 12:05:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 16 05:36:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	331748	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	475161	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	394254	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	159093	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	136316	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
35) Dibromofluoromethane	7.89	113	131495	46.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.66%	
50) Toluene-d8	10.33	98	460676	43.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.26%	
62) 4-Bromofluorobenzene	12.62	95	148341	36.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.94%	
Target Compounds						
52) Toluene	10.39	92	52786	6.554	ug/l	97
68) m/p-Xylenes	11.84	106	18136	3.137	ug/l	100
69) o-Xylene	12.17	106	6522	1.184	ug/l	96
95) Naphthalene	15.38	128	68587	10.965	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW101618\
Data File : VW006118.D
Acq On : 17 Oct 2018 10:23
Operator : VA/AP
Sample : J5200-11
Misc : 6.70G/5ML/MSVOA W/SOIL
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
OR-01-101618-B

Quant Time: Oct 17 12:05:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 16 05:36:33 2018
Response via : Initial Calibration

